

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022730.D
 Acq On : 17 Jun 2021 11:09
 Operator : JC/MD
 Sample : M2750-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB01-20210615

Quant Time: Jun 18 01:15:59 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	60209	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	123240	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	109295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	42496	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	51524	47.582	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.160%
35) Dibromofluoromethane	5.397	113	37297	47.434	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.860%
50) Toluene-d8	8.653	98	150561	48.574	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.140%
62) 4-Bromofluorobenzene	11.085	95	51993	44.352	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.700%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022730.D
Acq On : 17 Jun 2021 11:09
Operator : JC/MD
Sample : M2750-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB01-20210615

Quant Time: Jun 18 01:15:59 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

