

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022736.D
 Acq On : 17 Jun 2021 13:30
 Operator : JC/MD
 Sample : M2750-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE123D2-20210616

Quant Time: Jun 18 01:16:52 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	56143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	113864	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	99174	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48551	48.084	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.160%
35) Dibromofluoromethane	5.397	113	34455	47.427	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.860%
50) Toluene-d8	8.653	98	138466	48.350	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.700%
62) 4-Bromofluorobenzene	11.085	95	46773	43.184	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.360%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1237	2.736	ug/l	95
44) Trichloroethene	7.129	130	2570	3.244	ug/l	98
64) Tetrachloroethene	9.275	164	1590	2.535	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022736.D
Acq On : 17 Jun 2021 13:30
Operator : JC/MD
Sample : M2750-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE123D2-20210616

Quant Time: Jun 18 01:16:52 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

