

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022738.D
 Acq On : 17 Jun 2021 14:17
 Operator : JC/MD
 Sample : M2750-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178S-20210616

Quant Time: Jun 18 01:17:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	57479	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	114446	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	103143	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38064	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	49059	47.458	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.920%
35) Dibromofluoromethane	5.397	113	35358	48.423	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.653	98	140568	48.834	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.660%
62) 4-Bromofluorobenzene	11.085	95	48107	44.190	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.380%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1251	2.703	ug/l	97
19) Methyl tert-butyl Ether	3.123	73	4007	1.655	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022738.D
Acq On : 17 Jun 2021 14:17
Operator : JC/MD
Sample : M2750-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178S-20210616

Quant Time: Jun 18 01:17:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

