

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022740.D
 Acq On : 17 Jun 2021 15:04
 Operator : JC/MD
 Sample : M2750-19
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW178D-20210616

Quant Time: Jun 18 01:17:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55451	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	108211	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97333	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36821	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47226	47.355	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	94.720%
35) Dibromofluoromethane	5.397	113	33431	48.422	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.840%
50) Toluene-d8	8.653	98	134608	49.458	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.920%
62) 4-Bromofluorobenzene	11.085	95	45945	44.636	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.280%
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.319	96	2330	3.461	ug/l	90
16) Acetone	2.392	43	1309	2.931	ug/l #	82
27) cis-1,2-Dichloroethene	4.501	96	44970	49.477	ug/l	93
44) Trichloroethene	7.135	130	370496	492.019	ug/l	92
45) 1,2-Dichloropropane	7.434	63	37665	43.519	ug/l	95
64) Tetrachloroethene	9.275	164	17549	28.511	ug/l	94
76) 1,2,3-Trichloropropane	11.238	75	4344	4.450	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022740.D
Acq On : 17 Jun 2021 15:04
Operator : JC/MD
Sample : M2750-19
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW178D-20210616

Quant Time: Jun 18 01:17:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

