

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022744.D
 Acq On : 17 Jun 2021 16:39
 Operator : JC/MD
 Sample : M2750-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D1-20210615

Quant Time: Jun 18 01:18:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55640	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	112583	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	100872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38632	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	48383	48.351	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.700%
35) Dibromofluoromethane	5.397	113	34292	47.740	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.480%
50) Toluene-d8	8.653	98	138355	48.861	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.720%
62) 4-Bromofluorobenzene	11.085	95	47624	44.470	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.940%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	2.331	101	7471	10.970	ug/l	87
12) 1,1-Dichloroethene	2.325	96	1711	2.533	ug/l	# 86
16) Acetone	2.392	43	859	1.917	ug/l	# 83
19) Methyl tert-butyl Ether	3.123	73	1657	0.707	ug/l	93
27) cis-1,2-Dichloroethene	4.501	96	4233	4.641	ug/l	90
30) Chloroform	5.093	83	1227	0.821	ug/l	93
44) Trichloroethene	7.135	130	219634	280.347	ug/l	93
64) Tetrachloroethene	9.281	164	7103	11.135	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022744.D
Acq On : 17 Jun 2021 16:39
Operator : JC/MD
Sample : M2750-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D1-20210615

Quant Time: Jun 18 01:18:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

