

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022746.D
 Acq On : 17 Jun 2021 17:26
 Operator : JC/MD
 Sample : M2750-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE125D3-20210615

Quant Time: Jun 18 01:18:22 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53248	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	110433	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97400	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35747	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47856	49.972	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.940%
35) Dibromofluoromethane	5.397	113	33058	46.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.840%
50) Toluene-d8	8.653	98	135496	48.783	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.560%
62) 4-Bromofluorobenzene	11.085	95	46264	44.041	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.080%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	61171	93.851	ug/l	89
12) 1,1-Dichloroethene	2.331	96	718	1.111	ug/l	# 60
16) Acetone	2.386	43	932	2.173	ug/l	90
27) cis-1,2-Dichloroethene	4.501	96	2321	2.659	ug/l	94
30) Chloroform	5.111	83	612	0.428	ug/l	# 71
44) Trichloroethene	7.135	130	190732	248.196	ug/l	94
64) Tetrachloroethene	9.275	164	5569	9.041	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022746.D
Acq On : 17 Jun 2021 17:26
Operator : JC/MD
Sample : M2750-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE125D3-20210615

Quant Time: Jun 18 01:18:22 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

