

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022748.D
 Acq On : 17 Jun 2021 18:13
 Operator : JC/MD
 Sample : M2750-21
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW179D1-20210616

Quant Time: Jun 18 01:18:40 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54465	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109839	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	36405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	46719	47.695	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	95.380%
35) Dibromofluoromethane	5.397	113	33145	47.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.600%
50) Toluene-d8	8.653	98	136202	49.302	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.085	95	46203	44.221	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.440%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	1376	3.137	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	1500	0.654	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	16652	18.653	ug/l	93
44) Trichloroethene	7.135	130	133048	174.069	ug/l	91
64) Tetrachloroethene	9.281	164	9568	15.537	ug/l	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022748.D
Acq On : 17 Jun 2021 18:13
Operator : JC/MD
Sample : M2750-21
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW179D1-20210616

Quant Time: Jun 18 01:18:40 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

