

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
 Data File : VX022752.D
 Acq On : 17 Jun 2021 19:47
 Operator : JC/MD
 Sample : M2750-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D1-20210615

Manual Integrations
 APPROVED

MMDadoda
 6/18/2021 8:59:35 AM

Quant Time: Jun 18 01:19:34 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	61881	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	121924	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	111525	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	41559	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52919	47.550	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	95.100%		
35) Dibromofluoromethane	5.397	113	37397	48.074	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.140%		
50) Toluene-d8	8.653	98	152962	49.881	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	99.760%		
62) 4-Bromofluorobenzene	11.085	95	52899	45.612	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	91.220%		
Target Compounds						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.331	101	10801	14.259	ug/l	92
12) 1,1-Dichloroethene	2.325	96	5738	7.638	ug/l	96
16) Acetone	2.386	43	1126	2.260	ug/l	90
27) cis-1,2-Dichloroethene	4.501	96	3491	3.442	ug/l	89
30) Chloroform	5.099	83	987m	0.594	ug/l	
44) Trichloroethene	7.135	130	706195	832.348	ug/l	89
64) Tetrachloroethene	9.275	164	3247	4.604	ug/l #	78
95) Naphthalene	13.780	128	977	1.743	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061721\
Data File : VX022752.D
Acq On : 17 Jun 2021 19:47
Operator : JC/MD
Sample : M2750-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D1-20210615

Manual Integrations
APPROVED

MMDadoda
6/18/2021 8:59:35 AM

Quant Time: Jun 18 01:19:34 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

