

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061724\
 Data File : VX041908.D
 Acq On : 17 Jun 2024 14:41
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 18 02:08:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Sat Jun 15 04:02:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	4.898	128	29336	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	6.763	114	146035	30.000	ug/l	0.00
57) Chlorobenzene-d5	10.055	117	132491	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	5.952	65	59689	29.557	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	98.533%
60) 4-Bromofluorobenzene	11.079	95	56022	27.316	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	91.067%
63) Toluene-d8	8.647	98	166631	29.396	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061724\
Data File : VX041908.D
Acq On : 17 Jun 2024 14:41
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 18 02:08:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X061424W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Sat Jun 15 04:02:59 2024
Response via : Initial Calibration

