

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
 Data File : VX046723.D
 Acq On : 17 Jun 2025 15:23
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Jun 18 03:34:59 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	117985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	190185	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	165765	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	73855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	73294	44.912	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	89.820%
35) Dibromofluoromethane	5.397	113	60232	47.867	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.740%
50) Toluene-d8	8.647	98	225712	49.927	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.860%
62) 4-Bromofluorobenzene	11.079	95	79570	47.216	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.440%
Target Compounds						
					Qvalue	
17) Carbon Disulfide	2.532	76	5142	1.307	ug/l	100
93) 1,2,4-Trichlorobenzene	13.591	180	2718	1.603	ug/l	92
94) Hexachlorobutadiene	13.725	225	1238	1.736	ug/l	88
96) 1,2,3-Trichlorobenzene	13.963	180	2255	1.390	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061725\
Data File : VX046723.D
Acq On : 17 Jun 2025 15:23
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Jun 18 03:34:59 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

