

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061819\
 Data File : VX010335.D
 Acq On : 18 Jun 2019 12:06
 Operator : JC/SP
 Sample : K3427-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-190617

Quant Time: Jun 19 01:30:33 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
 Quant Title : SW846 8260
 QLast Update : Sat Jun 08 02:12:13 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	310717	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	461985	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	414599	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	191200	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	138182	45.56	ug/l	0.00
Spiked Amount			Recovery	=	91.12%	
35) Dibromofluoromethane	5.50	113	139269	48.87	ug/l	0.00
Spiked Amount			Recovery	=	97.74%	
50) Toluene-d8	8.71	98	543547	50.89	ug/l	0.00
Spiked Amount			Recovery	=	101.78%	
62) 4-Bromofluorobenzene	11.13	95	182767	45.82	ug/l	0.00
Spiked Amount			Recovery	=	91.64%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061819\
Data File : VX010335.D
Acq On : 18 Jun 2019 12:06
Operator : JC/SP
Sample : K3427-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-190617

Quant Time: Jun 19 01:30:33 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X060719W.M
Quant Title : SW846 8260
QLast Update : Sat Jun 08 02:12:13 2019
Response via : Initial Calibration

