


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016866.D
 Acq On : 19 Jun 2020 02:33
 Operator : JC/SP
 Sample : L3046-15
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ISS-C-C

Quant Time: Jun 19 04:05:17 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	284576	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	394946	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	410776	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	225561	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	152580	44.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	89.16%	
35) Dibromofluoromethane	5.47	113	145995	57.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	115.58%	
50) Toluene-d8	8.70	98	478268	47.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.64%	
62) 4-Bromofluorobenzene	11.12	95	197595	51.02	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.04%	
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.18	85	185	Below Cal	#	43
5) Bromomethane	1.64	94	271	Below Cal	#	73
11) Tert butyl alcohol	3.00	59	1115	1.927	ug/l #	82
13) Acrolein	2.28	56	356	1.015	ug/l #	16
16) Acetone	2.42	43	13231	Below Cal		96
18) Methyl Acetate	2.75	43	2315	0.733	ug/l	90
20) Methylene Chloride	2.84	84	6147	2.063	ug/l	96
25) 2-Butanone	4.65	43	6227	3.642	ug/l #	85
29) Tetrahydrofuran	5.11	42	840	0.780	ug/l #	61
31) Cyclohexane	5.63	56	5187	1.226	ug/l #	7
36) 1,1-Dichloropropene	5.64	75	19851	5.444	ug/l #	48
38) Carbon Tetrachloride	5.63	117	26851	6.562	ug/l #	15
43) Isopropyl Acetate	6.42	43	44971	7.993	ug/l	98
48) Methyl methacrylate	7.65	41	3344	1.262	ug/l #	12
51) 4-Methyl-2-Pentanone	8.59	43	37285	11.017	ug/l #	44
52) Toluene	8.77	92	1483	0.204	ug/l	89
54) cis-1,3-Dichloropropene	8.59	75	18571	3.980	ug/l #	62
56) Ethyl methacrylate	9.29	69	859	0.217	ug/l #	1
57) 1,3-Dichloropropane	9.52	76	3333	0.717	ug/l #	42
59) 2-Hexanone	9.42	43	37130	5.329	ug/l #	21
73) Isopropylbenzene	11.00	105	404	Below Cal		94
74) N-amyl acetate	10.80	43	2326	0.459	ug/l #	43
76) 1,2,3-Trichloropropane	11.12	75	95108	22.854	ug/l #	37
80) 1,3,5-Trimethylbenzene	11.49	105	662	0.257	ug/l	74
81) trans-1,4-Dichloro-2-buten	11.12	75	95108	56.537	ug/l #	14
89) n-Butylbenzene	12.37	91	889	3.081	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.64 min Scan# 746

Delta R.T. 0.01 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :


ISS-C-C

Tot Ion:168 Resp: 284576

Ion Ratio Lower Upper

168 100

99 51.4 42.2 63.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.64

100000

80000

60000

40000

20000



0

5.50

5.60

5.70

Time-->



#2

Dichlorodifluoromethane

Concen: Below Cal

RT: 1.18 min Scan# 16

Delta R.T. 0.00 min

Lab File: VX016866.D


Acq: 19 Jun 2020 02:33

Tgt Ion: 85 Resp: 185

Ion Ratio Lower Upper

85 100

87 0.0 16.0 47.9#



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.18

250

200

150

100

50

0


1.16


1.18

1.20

1.22

Time-->





#5

Bromomethane

Concen: Below Cal

RT: 1.64 min Scan# 91

Delta R.T. 0.02 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :


ISS-C-C

Tot Ion: 94 Resp: 271

Ion Ratio Lower Upper




94 100

96 67.5 74.4 111.6#



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#11

Tert butyl alcohol

Concen: 1.927 ug/l

RT: 3.00 min Scan# 314

Delta R.T. -0.01 min

Lab File: VX016866.D


Acq: 19 Jun 2020 02:33

Tgt Ion: 59 Resp: 1115

Ion Ratio Lower Upper



59 100


57 3.7 8.4 12.6#



Abundance Ion 59.00 (58.70 to 59.70): VX0

Ion 57.00 (56.70 to 57.70): VX0









#13
Acrolein
Concen: 1.015 ug/l
RT: 2.28 min Scan# 195
Delta R.T. 0.01 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33




Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C


Tot Ion: 56 Resp: 356
Ion Ratio Lower Upper
56 100
55 137.9 55.7 83.5#



#16
Acetone
Concen: Below Cal
RT: 2.42 min Scan# 218
Delta R.T. 0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion: 43 Resp: 13231
Ion Ratio Lower Upper
43 100
58 30.9 26.6 39.8









#18
Methyl Acetate
Concen: 0.733 ug/l
RT: 2.75 min Scan# 272
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33




Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C


Tot Ion: 43 Resp: 2315
Ion Ratio Lower Upper
43 100
74 22.2 21.8 32.8



#20
Methylene Chloride
Concen: 2.063 ug/l
RT: 2.84 min Scan# 287
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion: 84 Resp: 6147
Ion Ratio Lower Upper
84 100
49 109.2 86.6 129.8
51 28.8 27.1 40.7
86 62.5 53.1 79.7









#25
2-Butanone
Concen: 3.642 ug/l
RT: 4.65 min Scan# 584
Delta R.T. 0.02 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33




Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C


Tot Ion: 43 Resp: 6227
Ion Ratio Lower Upper
43 100
72 37.5 23.5 35.3#



#29
Tetrahydrofuran
Concen: 0.780 ug/l
RT: 5.11 min Scan# 660
Delta R.T. 0.02 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion: 42 Resp: 840
Ion Ratio Lower Upper
42 100
72 30.0 41.1 61.7#
71 15.5 38.4 57.6#









#31
Cyclohexane
Concen: 1.226 ug/l
RT: 5.63 min Scan# 745
Delta R.T. 0.08 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C

Tot Ion: 56 Resp: 5187
Ion Ratio Lower Upper
56 100
69 178.6 26.1 39.1#
84 122.6 73.0 109.6#




Abundance Ion 56.00 (55.70 to 56.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 84.00 (83.70 to 84.70): VX0






#33
1,2-Dichloroethane-d4
Concen: 44.581 ug/l
RT: 6.03 min Scan# 811
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion: 65 Resp: 152580
Ion Ratio Lower Upper
65 100
67 51.3 0.0 102.6



Abundance Ion 64.90 (64.60 to 65.60): VX0
Ion 66.90 (66.60 to 67.60): VX0





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :

ISS-C-C


Tot Ion:114 Resp: 394946

Ion Ratio Lower Upper

114 100

63 19.4 0.0 37.0




88 16.2 0.0 32.0



Abundance Ion 113.90 (113.60 to 114.60): V

Ion 63.00 (62.70 to 63.70): VX0

Ion 87.90 (87.60 to 88.60): VX0



#35

Dibromofluoromethane

Concen: 57.786 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.01 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33


Tgt Ion:113 Resp: 145995

Ion Ratio Lower Upper

113 100

111 102.9 81.0 121.4



192 22.9 18.5 27.7




Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V





#36

1,1-Dichloropropene

Concen: 5.444 ug/l

RT: 5.64 min Scan# 746

Delta R.T. -0.13 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :

ISS-C-C


Tot Ion: 75 Resp: 19851

Ion Ratio Lower Upper

75 100

110 9.3 18.9 56.9#




77 0.0 25.6 38.4#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0



#38

Carbon Tetrachloride

Concen: 6.562 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.13 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33


Tgt Ion: 117 Resp: 26851

Ion Ratio Lower Upper

117 100

119 4.7 77.4 116.2#



121 0.0 24.7 37.1#




Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V





#43

Isopropyl Acetate

Concen: 7.993 ug/l

RT: 6.42 min Scan# 874

Delta R.T. 0.01 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :

ISS-C-C


Tot Ion: 43 Resp: 44971

Ion Ratio Lower Upper

43 100

61 21.6 17.9 26.9




87 14.7 12.4 18.6



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 61.00 (60.70 to 61.70): VX0

Ion 87.00 (86.70 to 87.70): VX0



#48

Methyl methacrylate

Concen: 1.262 ug/l

RT: 7.65 min Scan# 1077

Delta R.T. -0.09 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33


Tgt Ion: 41 Resp: 3344

Ion Ratio Lower Upper

41 100

69 0.0 73.3 109.9#



39 0.0 42.4 63.6#




Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 69.00 (68.70 to 69.70): VX0

Ion 39.00 (38.70 to 39.70): VX0









#50
Toluene-d8
Concen: 47.315 ug/l
RT: 8.70 min Scan# 1248
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33




Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C


Tot Ion: 98 Resp: 478268
Ion Ratio Lower Upper
98 100
100 66.6 54.0 81.0



#51
4-Methyl-2-Pentanone
Concen: 11.017 ug/l
RT: 8.59 min Scan# 1231
Delta R.T. -0.03 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion: 43 Resp: 37285
Ion Ratio Lower Upper
43 100
58 7.3 34.3 51.5#





#52

Toluene

Concen: 0.204 ug/l

RT: 8.77 min Scan# 1260

Delta R.T. 0.01 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :





ISS-C-C

Tot Ion: 92 Resp: 1483

Ion Ratio Lower Upper

92 100

91 183.3 134.2 201.2



#54

cis-1,3-Dichloropropene

Concen: 3.980 ug/l

RT: 8.59 min Scan# 1231

Delta R.T. 0.18 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33




Tgt Ion: 75 Resp: 18571


Ion Ratio Lower Upper

75 100

77 0.6 24.9 37.3#

39 26.4 33.4 50.0#





#56

Ethyl methacrylate

Concen: 0.217 ug/l

RT: 9.29 min Scan# 1345

Delta R.T. 0.13 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :

ISS-C-C


Tot Ion: 69 Resp: 859

Ion Ratio Lower Upper

69 100

41 3364.8 47.9 71.9#




39 1335.2 24.2 36.4#



Abundance Ion 69.00 (68.70 to 69.70): VX0

Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0



#57

1,3-Dichloropropane

Concen: 0.717 ug/l

RT: 9.52 min Scan# 1384

Delta R.T. 0.17 min

Lab File: VX016866.D


Acq: 19 Jun 2020 02:33

Tgt Ion: 76 Resp: 3333

Ion Ratio Lower Upper



76 100


78 0.0 25.8 38.6#



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0









#59
2-Hexanone
Concen: 5.329 ug/l
RT: 9.42 min Scan# 1367
Delta R.T. -0.05 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33




Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C


Tot Ion: 43 Resp: 37130
Ion Ratio Lower Upper
43 100
58 0.0 29.8 89.4#



#62
4-Bromofluorobenzene
Concen: 51.021 ug/l
RT: 11.12 min Scan# 1646
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion: 95 Resp: 197595
Ion Ratio Lower Upper
95 100
174 86.8 0.0 174.2
176 84.5 0.0 170.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.10 min Scan# 1478
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Instrument :
MSVOA_X
ClientSampleId :
ISS-C-C

Tot Ion:117 Resp: 410776
Ion Ratio Lower Upper
117 100
82 54.4 43.4 65.0
119 32.9 25.7 38.5




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V






#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX016866.D
Acq: 19 Jun 2020 02:33

Tgt Ion:152 Resp: 225561
Ion Ratio Lower Upper
152 100
115 57.1 40.9 122.8
150 154.2 0.0 344.2



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V





#73

Isopropylbenzene

Concen: Below Cal

RT: 11.00 min Scan# 1626

Delta R.T. -0.00 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :


ISS-C-C

Tot Ion:105 Resp: 404

Ion Ratio Lower Upper




105 100

120 23.5 13.2 39.6



Abundance Ion 105.00 (104.70 to 105.70): V

Ion 120.00 (119.70 to 120.70): V



#74

N-ethyl acetate

Concen: 0.459 ug/l

RT: 10.80 min Scan# 1594

Delta R.T. -0.07 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Tgt Ion: 43 Resp: 2326


Ion Ratio Lower Upper

43 100

70 0.0 39.9 59.9#

55 9.8 21.8 32.6#

61 0.0 22.2 33.4#






Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 70.00 (69.70 to 70.70): VX0

Ion 55.00 (54.70 to 55.70): VX0

Ion 61.00 (60.70 to 61.70): VX0









#76
 1,2,3-Trichloropropane
 Concen: 22.854 ug/l
 RT: 11.12 min Scan# 1646
 Delta R.T. -0.16 min
 Lab File: VX016866.D
 Acq: 19 Jun 2020 02:33




Instrument :
 MSVOA_X
 ClientSampleId :
 ISS-C-C


Tot Ion: 75 Resp: 95108
 Ion Ratio Lower Upper
 75 100
 77 1.4 20.5 61.5#



#80
 1,3,5-Trimethylbenzene
 Concen: 0.257 ug/l
 RT: 11.49 min Scan# 1706
 Delta R.T. 0.00 min
 Lab File: VX016866.D
 Acq: 19 Jun 2020 02:33

Tgt Ion: 105 Resp: 662
 Ion Ratio Lower Upper
 105 100
 120 67.2 24.6 74.0





#81

trans-1,4-Dichloro-2-butene

Concen: 56.537 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. 0.06 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33

Instrument :

MSVOA_X

ClientSampleId :

ISS-C-C


Tot Ion: 75 Resp: 95108

Ion Ratio Lower Upper

75 100

53 0.1 69.8 104.8#




89 0.0 38.3 57.5#



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 53.00 (52.70 to 53.70): VX0

Ion 88.90 (88.60 to 89.60): VX0



#89

n-Butylbenzene

Concen: 3.081 ug/l

RT: 12.37 min Scan# 1851

Delta R.T. 0.00 min

Lab File: VX016866.D

Acq: 19 Jun 2020 02:33


Tgt Ion: 91 Resp: 889

Ion Ratio Lower Upper

91 100

92 61.0 27.4 82.2

134 30.7 13.9 41.6



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 92.00 (91.70 to 92.70): VX0

Ion 134.00 (133.70 to 134.70): V

