

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016851.D
 Acq On : 18 Jun 2020 20:53
 Operator : JC/SP
 Sample : L3020-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01_061520

Quant Time: Jun 19 03:28:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	267406	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	403677	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	402160	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	215266	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	142012	44.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.32%	
35) Dibromofluoromethane	5.47	113	138599	53.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.34%	
50) Toluene-d8	8.70	98	507447	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
62) 4-Bromofluorobenzene	11.12	95	200292	50.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.20%	
Target Compounds						
89) n-Butylbenzene	12.37	91	1738	3.149	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061820\
Data File : VX016851.D
Acq On : 18 Jun 2020 20:53
Operator : JC/SP
Sample : L3020-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01_061520

Quant Time: Jun 19 03:28:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 18 15:20:47 2020
Response via : Initial Calibration

