

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016854.D
 Acq On : 18 Jun 2020 22:01
 Operator : JC/SP
 Sample : PB129596ZHE#14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129596ZHE#14

Quant Time: Jun 19 03:37:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	242320	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	467698	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	494048	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	220913	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	142711	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
35) Dibromofluoromethane	5.46	113	126749	42.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	84.72%	
50) Toluene-d8	8.70	98	489468	40.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	81.78%	
62) 4-Bromofluorobenzene	11.12	95	204918	44.68	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.36%	
Target Compounds						
20) Methylene Chloride	2.84	84	2871	1.132	ug/l	90
43) Isopropyl Acetate	6.41	43	75184	11.284	ug/l	99
89) n-Butylbenzene	12.37	91	1683	3.141	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061820\
Data File : VX016854.D
Acq On : 18 Jun 2020 22:01
Operator : JC/SP
Sample : PB129596ZHE#14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129596ZHE#14

Quant Time: Jun 19 03:37:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 18 15:20:47 2020
Response via : Initial Calibration

