

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016856.D
 Acq On : 18 Jun 2020 22:46
 Operator : JC/SP
 Sample : PB129596ZHE#16
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB129596ZHE#16

Quant Time: Jun 19 03:41:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	290375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	423253	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	431389	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	231421	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	143251	41.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	82.04%	
35) Dibromofluoromethane	5.46	113	157087	58.02	ug/l	0.00
Spiked Amount 50.000			Recovery	=	116.04%	
50) Toluene-d8	8.70	98	500804	46.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.46%	
62) 4-Bromofluorobenzene	11.12	95	225799	54.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.80%	
Target Compounds						
20) Methylene Chloride	2.84	84	3355	1.104	ug/l	93
43) Isopropyl Acetate	6.42	43	56091	9.302	ug/l	99
89) n-Butylbenzene	12.37	91	1582	3.128	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061820\
Data File : VX016856.D
Acq On : 18 Jun 2020 22:46
Operator : JC/SP
Sample : PB129596ZHE#16
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB129596ZHE#16

Quant Time: Jun 19 03:41:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 18 15:20:47 2020
Response via : Initial Calibration

