

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016859.D
 Acq On : 18 Jun 2020 23:55
 Operator : JC/SP
 Sample : L3020-11
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-22

Quant Time: Jun 19 03:49:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	238990	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	418631	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	382391	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	235861	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	139472	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
35) Dibromofluoromethane	5.46	113	130742	48.82	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.64%	
50) Toluene-d8	8.70	98	531501	49.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.22%	
62) 4-Bromofluorobenzene	11.12	95	194030	47.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.54%	
Target Compounds						
3) Chloromethane	1.31	50	5145	1.730	ug/l	98
30) Chloroform	5.18	83	3875	0.908	ug/l	89
47) Bromodichloromethane	7.88	83	3681	0.875	ug/l	91
60) Dibromochloromethane	9.57	129	2441	0.700	ug/l	95
64) Tetrachloroethene	9.32	164	2525	0.713	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061820\
Data File : VX016859.D
Acq On : 18 Jun 2020 23:55
Operator : JC/SP
Sample : L3020-11
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-22

Quant Time: Jun 19 03:49:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 18 15:20:47 2020
Response via : Initial Calibration

