

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016862.D
 Acq On : 19 Jun 2020 01:03
 Operator : JC/SP
 Sample : L3020-20
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-23

Quant Time: Jun 19 03:56:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	228227	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	454077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	434852	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	225286	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	135238	49.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.54%	
35) Dibromofluoromethane	5.46	113	116273	40.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	80.06%	
50) Toluene-d8	8.70	98	451724	38.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	77.74%	
62) 4-Bromofluorobenzene	11.12	95	204274	45.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.76%	
Target Compounds						
					Qvalue	
3) Chloromethane	1.31	50	3156	1.111	ug/l	97
30) Chloroform	5.18	83	3025	0.742	ug/l	92
47) Bromodichloromethane	7.87	83	2641	0.579	ug/l	99
60) Dibromochloromethane	9.56	129	2262	0.598	ug/l	97
68) m/p-Xylenes	10.34	106	2022	0.350	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061820\
Data File : VX016862.D
Acq On : 19 Jun 2020 01:03
Operator : JC/SP
Sample : L3020-20
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-23

Quant Time: Jun 19 03:56:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 18 15:20:47 2020
Response via : Initial Calibration

