

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061820\
 Data File : VX016864.D
 Acq On : 19 Jun 2020 01:48
 Operator : JC/SP
 Sample : L3046-13
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 33 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ISS-A-C

Quant Time: Jun 19 04:01:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 18 15:20:47 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	240713	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	368457	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	481729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	241810	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	137204	47.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.78%	
35) Dibromofluoromethane	5.47	113	141750	60.14	ug/l	0.00
Spiked Amount 50.000			Recovery	=	120.28%	
50) Toluene-d8	8.70	98	499372	52.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.90%	
62) 4-Bromofluorobenzene	11.12	95	237222	65.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	131.32%	
Target Compounds						
20) Methylene Chloride	2.84	84	6323	2.509	ug/l	95
43) Isopropyl Acetate	6.42	43	41440	7.895	ug/l	99
89) n-Butylbenzene	12.37	91	1168	3.095	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061820\
Data File : VX016864.D
Acq On : 19 Jun 2020 01:48
Operator : JC/SP
Sample : L3046-13
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 33 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ISS-A-C

Quant Time: Jun 19 04:01:58 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 18 15:20:47 2020
Response via : Initial Calibration

