


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX061820\
Data File : VX016837.D
Acq On : 18 Jun 2020 10:01
Operator : JC/SP
Sample : BFB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Title : SW846 8260
Last Update : Fri Jun 19 12:07:48 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1638

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	15.7	12427	PASS
75	95	30	60	46.8	37117	PASS
95	95	100	100	100.0	79325	PASS
96	95	5	9	6.6	5245	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	100	86.7	68738	PASS
175	174	5	9	7.6	5241	PASS
176	174	95	101	97.9	67304	PASS
177	176	5	9	6.1	4107	PASS