

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
 Data File : VX022769.D
 Acq On : 18 Jun 2021 15:48
 Operator : JC/MD
 Sample : M2750-09DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RE120D2-20210615DL

Quant Time: Jun 19 01:45:11 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	53129	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	106365	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	98961	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	38566	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	48756	51.026	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.060%
35) Dibromofluoromethane	5.397	113	33866	49.903	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.800%
50) Toluene-d8	8.653	98	140892	52.666	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.340%
62) 4-Bromofluorobenzene	11.085	95	50972	50.379	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.760%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.313	101	1858	2.857	ug/l	96
12) 1,1-Dichloroethene	2.313	96	188	0.291	ug/l #	67
44) Trichloroethene	7.129	130	54581	73.742	ug/l	91
64) Tetrachloroethene	9.275	164	708	1.131	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
Data File : VX022769.D
Acq On : 18 Jun 2021 15:48
Operator : JC/MD
Sample : M2750-09DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RE120D2-20210615DL

Quant Time: Jun 19 01:45:11 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

