

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
 Data File : VX022773.D
 Acq On : 18 Jun 2021 17:45
 Operator : JC/MD
 Sample : M2753-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 DUP02-20210616

Quant Time: Jun 19 01:46:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52231	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	107116	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	94131	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	34522	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	51040	54.335	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.660%	
35) Dibromofluoromethane	5.397	113	35459	51.884	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.760%	
50) Toluene-d8	8.653	98	141843	52.649	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 105.300%	
62) 4-Bromofluorobenzene	11.085	95	46577	45.712	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 91.420%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	1093	2.599	ug/l	# 82
19) Methyl tert-butyl Ether	3.117	73	1375	0.625	ug/l	93
27) cis-1,2-Dichloroethene	4.495	96	15927	18.604	ug/l	95
44) Trichloroethene	7.135	130	130175	174.640	ug/l	94
64) Tetrachloroethene	9.275	164	9204	15.462	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
Data File : VX022773.D
Acq On : 18 Jun 2021 17:45
Operator : JC/MD
Sample : M2753-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
DUP02-20210616

Quant Time: Jun 19 01:46:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

