

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
 Data File : VX022776.D
 Acq On : 18 Jun 2021 18:56
 Operator : JC/MD
 Sample : M2780-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 19 01:46:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	52143	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	104363	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	89568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	31207	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	45191	48.189	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.380%
35) Dibromofluoromethane	5.397	113	31762	47.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.400%
50) Toluene-d8	8.653	98	127231	48.472	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.940%
62) 4-Bromofluorobenzene	11.085	95	41201	41.503	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	83.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
Data File : VX022776.D
Acq On : 18 Jun 2021 18:56
Operator : JC/MD
Sample : M2780-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jun 19 01:46:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

