

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
 Data File : VX022778.D
 Acq On : 18 Jun 2021 19:43
 Operator : JC/MD
 Sample : M2748-17
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 FB

Quant Time: Jun 19 01:47:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	55288	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	110136	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97460	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35880	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	52293	52.591	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.180%
35) Dibromofluoromethane	5.391	113	37548	53.434	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.860%
50) Toluene-d8	8.653	98	148221	53.508	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.020%
62) 4-Bromofluorobenzene	11.085	95	49182	46.946	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.900%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3701	8.312	ug/l #	86
20) Methylene Chloride	2.794	84	1646	1.646	ug/l #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
Data File : VX022778.D
Acq On : 18 Jun 2021 19:43
Operator : JC/MD
Sample : M2748-17
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
FB

Quant Time: Jun 19 01:47:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration

