

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
 Data File : VX022774.D
 Acq On : 18 Jun 2021 18:09
 Operator : JC/MD
 Sample : M2780-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 19 01:46:32 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jun 10 12:20:00 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.562	168	54260	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.769	114	109641	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	97519	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	35451	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	47868	49.053	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.100%
35) Dibromofluoromethane	5.397	113	33422	47.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.560%
50) Toluene-d8	8.653	98	134913	48.924	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	97.840%
62) 4-Bromofluorobenzene	11.085	95	44845	42.999	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	86.000%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
Data File : VX022774.D
Acq On : 18 Jun 2021 18:09
Operator : JC/MD
Sample : M2780-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Quant Time: Jun 19 01:46:32 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
Quant Title : SW846 8260
QLast Update : Thu Jun 10 12:20:00 2021
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.562 min Scan# 734
Delta R.T. -0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:168 Resp: 54260
Ion Ratio Lower Upper
168 100
99 72.0 45.8 68.6#



#33
1,2-Dichloroethane-d4
Concen: 49.053 ug/l
RT: 5.964 min Scan# 800
Delta R.T. 0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09

Tgt Ion: 65 Resp: 47868
Ion Ratio Lower Upper
65 100
67 51.9 0.0 105.4









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 6.769 min Scan# 932
Delta R.T. 0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion:114 Resp: 109641
Ion Ratio Lower Upper
114 100
63 23.0 0.0 40.8
88 16.5 0.0 32.8



#35
Dibromofluoromethane
Concen: 47.777 ug/l
RT: 5.397 min Scan# 707
Delta R.T. 0.006 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09

Tgt Ion:113 Resp: 33422
Ion Ratio Lower Upper
113 100
111 103.8 81.9 122.9
192 15.3 16.7 25.1#









#50
Toluene-d8
Concen: 48.924 ug/l
RT: 8.653 min Scan# 1241
Delta R.T. -0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09




Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6


Tgt Ion: 98 Resp: 134913
Ion Ratio Lower Upper
98 100
100 66.4 51.9 77.9



#62
4-Bromofluorobenzene
Concen: 42.999 ug/l
RT: 11.085 min Scan# 1640
Delta R.T. 0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09

Tgt Ion: 95 Resp: 44845
Ion Ratio Lower Upper
95 100
174 62.1 0.0 154.6
176 61.8 0.0 155.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. 0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SOIL-REF-6

Tgt Ion:117	Resp:	97519
Ion Ratio	Lower	Upper
117	100	
82	60.9	47.4 71.2
119	31.3	26.6 39.8



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX022774.D
Acq: 18 Jun 2021 18:09

Tgt Ion:152	Resp:	35451
Ion Ratio	Lower	Upper
152	100	
115	68.5	41.1 123.3
150	161.4	0.0 349.0

