



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061821\
 Data File : VX022754.D
 Acq On : 18 Jun 2021 08:47
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061021W.M
 Title : SW846 8260
 Last Update : Thu Jun 10 12:20:00 2021



AutoFind: Scans 1639, 1640, 1641; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	20.8	4394	PASS
75	95	30	60	57.7	12210	PASS
95	95	100	100	100.0	21156	PASS
96	95	5	9	6.8	1440	PASS
173	174	0.00	2	0.6	79	PASS
174	95	50	100	63.5	13432	PASS
175	174	5	9	6.7	905	PASS
176	174	95	101	96.1	12909	PASS
177	176	5	9	6.4	830	PASS