

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041917.D
 Acq On : 18 Jun 2024 11:50
 Operator : JC/MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VIBLK

Quant Time: Jun 19 07:52:40 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:27:13 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	151229	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	232646	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229917	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	117406	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75269	43.162	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.320%
35) Dibromofluoromethane	5.391	113	71640	43.399	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.800%
50) Toluene-d8	8.647	98	272464	46.983	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	93.960%
62) 4-Bromofluorobenzene	11.079	95	104744	48.998	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.000%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041917.D
Acq On : 18 Jun 2024 11:50
Operator : JC/MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VIBLK

Quant Time: Jun 19 07:52:40 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:27:13 2024
Response via : Initial Calibration

