

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041925.D
 Acq On : 18 Jun 2024 16:00
 Operator : JC/MD
 Sample : P2928-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-TB-20240613

Quant Time: Jun 19 08:12:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	134272	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	205991	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	208554	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	106820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66362	42.860	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	85.720%
35) Dibromofluoromethane	5.385	113	63807	43.655	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.320%
50) Toluene-d8	8.653	98	245853	47.880	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.760%
62) 4-Bromofluorobenzene	11.085	95	92961	49.113	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.220%
Target Compounds						
16) Acetone	2.392	43	2313	3.402	ug/l	95
20) Methylene Chloride	2.776	84	1522	1.022	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041925.D
Acq On : 18 Jun 2024 16:00
Operator : JC/MD
Sample : P2928-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-TB-20240613

Quant Time: Jun 19 08:12:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

