

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041927.D
 Acq On : 18 Jun 2024 16:47
 Operator : JC/MD
 Sample : P2928-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-GW-560-562

Quant Time: Jun 19 08:13:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135902	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	206392	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207441	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104341	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67947	43.358	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.720%
35) Dibromofluoromethane	5.391	113	63887	43.625	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.260%
50) Toluene-d8	8.653	98	243538	47.337	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.680%
62) 4-Bromofluorobenzene	11.079	95	91918	48.468	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.940%
Target Compounds						
16) Acetone	2.392	43	5269	7.657	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041927.D
Acq On : 18 Jun 2024 16:47
Operator : JC/MD
Sample : P2928-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-GW-560-562

Quant Time: Jun 19 08:13:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

