

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041928.D
 Acq On : 18 Jun 2024 17:11
 Operator : JC/MD
 Sample : P2927-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-458-460

Quant Time: Jun 19 08:13:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129911	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	198675	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	194807	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90009	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	64496	43.054	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.100%
35) Dibromofluoromethane	5.391	113	61170	43.392	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	86.780%
50) Toluene-d8	8.653	98	234672	47.386	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.780%
62) 4-Bromofluorobenzene	11.079	95	82345	45.107	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.220%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	724	0.645	ug/l	# 82
16) Acetone	2.392	43	2819	4.286	ug/l	# 87
20) Methylene Chloride	2.794	84	898	0.623	ug/l	# 76

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041928.D
Acq On : 18 Jun 2024 17:11
Operator : JC/MD
Sample : P2927-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-GW-458-460

Quant Time: Jun 19 08:13:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

