

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041930.D
 Acq On : 18 Jun 2024 17:59
 Operator : JC/MD
 Sample : P2927-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-478-480

Quant Time: Jun 19 08:48:35 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129863	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	195309	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97551	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64887	43.331	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	86.660%
35) Dibromofluoromethane	5.385	113	61076	44.072	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.140%
50) Toluene-d8	8.653	98	236066	48.489	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	96.980%
62) 4-Bromofluorobenzene	11.079	95	86297	48.086	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.180%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	2.319	101	11094	9.880	ug/l	87
12) 1,1-Dichloroethene	2.306	96	1509	1.347	ug/l	# 78
16) Acetone	2.392	43	6375	9.695	ug/l	91
27) cis-1,2-Dichloroethene	4.489	96	3877	2.723	ug/l	78
30) Chloroform	5.099	83	3418	1.560	ug/l	99
38) Carbon Tetrachloride	5.678	117	1277	0.785	ug/l	86
44) Trichloroethene	7.129	130	171277	128.742	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041930.D
Acq On : 18 Jun 2024 17:59
Operator : JC/MD
Sample : P2927-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-GW-478-480

Quant Time: Jun 19 08:48:35 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

