

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041931.D
 Acq On : 18 Jun 2024 18:22
 Operator : JC/MD
 Sample : P2927-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-189-GW-498-500

Quant Time: Jun 19 08:49:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132041	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	204686	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	207742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	104564	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67435	44.289	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	88.580%
35) Dibromofluoromethane	5.391	113	63940	44.025	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.060%
50) Toluene-d8	8.653	98	244506	47.922	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	95.840%
62) 4-Bromofluorobenzene	11.079	95	93226	49.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.140%
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.166	85	2704	2.688	ug/l	93
9) 1,1,2-Trichlorotrifluo...	2.319	101	18647	16.333	ug/l	86
12) 1,1-Dichloroethene	2.306	96	3499	3.072	ug/l	83
16) Acetone	2.392	43	4970	7.434	ug/l	99
17) Carbon Disulfide	2.501	76	8717	3.754	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	5387	3.721	ug/l	78
30) Chloroform	5.105	83	9263	4.159	ug/l	96
38) Carbon Tetrachloride	5.672	117	2815	1.652	ug/l	86
44) Trichloroethene	7.129	130	567167	406.787	ug/l	87
55) 1,1,2-Trichloroethane	9.159	97	1791	1.298	ug/l	96
95) Naphthalene	13.786	128	407	0.801	ug/l	# 69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041931.D
Acq On : 18 Jun 2024 18:22
Operator : JC/MD
Sample : P2927-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-189-GW-498-500

Quant Time: Jun 19 08:49:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

