

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041933.D
 Acq On : 18 Jun 2024 19:10
 Operator : JC/MD
 Sample : P2928-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BP-VPB-196-GW-540-542

Quant Time: Jun 19 08:50:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	128011	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	196368	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	196429	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95537	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66919	45.334	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	90.660%
35) Dibromofluoromethane	5.385	113	61931	44.448	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	88.900%
50) Toluene-d8	8.647	98	232124	47.422	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	94.840%
62) 4-Bromofluorobenzene	11.079	95	88232	48.899	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.800%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	5453	8.413	ug/l	# 83
17) Carbon Disulfide	2.495	76	2171	0.964	ug/l	# 94
44) Trichloroethene	7.135	130	2648	1.980	ug/l	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041933.D
Acq On : 18 Jun 2024 19:10
Operator : JC/MD
Sample : P2928-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BP-VPB-196-GW-540-542

Quant Time: Jun 19 08:50:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

