

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX061824\
 Data File : VX041934.D
 Acq On : 18 Jun 2024 19:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 19 08:50:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	50.000	50.118	-0.2	100	0.00
3 P	Chloromethane	50.000	48.546	2.9	98	0.00
4 C	Vinyl Chloride	50.000	49.686	0.6#	100	0.00
5 T	Bromomethane	50.000	48.679	2.6	97	0.00
6 T	Chloroethane	50.000	63.142	-26.3#	118	0.00
7 T	Trichlorofluoromethane	50.000	55.624	-11.2	116	0.00
8 T	Diethyl Ether	50.000	49.979	0.0	100	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	48.246	3.5	97	0.00
10 T	Methyl Iodide	50.000	51.269	-2.5	102	0.00
11 T	Tert butyl alcohol	250.000	276.721	-10.7	107	0.00
12 CM	1,1-Dichloroethene	50.000	48.679	2.6#	96	0.00
13 T	Acrolein	250.000	244.450	2.2	96	0.00
14 T	Allyl chloride	50.000	48.553	2.9	94	0.00
15 T	Acrylonitrile	250.000	270.108	-8.0	103	0.00
16 T	Acetone	250.000	259.054	-3.6	111	0.00
17 T	Carbon Disulfide	50.000	46.020	8.0	97	0.00
18 T	Methyl Acetate	50.000	54.630	-9.3	101	0.00
19 T	Methyl tert-butyl Ether	50.000	51.812	-3.6	100	0.00
20 T	Methylene Chloride	50.000	46.174	7.7	100	0.00
21 T	trans-1,2-Dichloroethene	50.000	49.522	1.0	99	0.00
22 T	Diisopropyl ether	50.000	51.675	-3.3	99	0.00
23 T	Vinyl Acetate	250.000	262.145	-4.9	98	0.00
24 P	1,1-Dichloroethane	50.000	51.219	-2.4	100	0.00
25 T	2-Butanone	250.000	271.182	-8.5	103	0.00
26 T	2,2-Dichloropropane	50.000	34.242	31.5#	67	0.00
27 T	cis-1,2-Dichloroethene	50.000	50.698	-1.4	98	0.00
28 T	Bromochloromethane	50.000	48.358	3.3	99	0.00
29 T	Tetrahydrofuran	250.000	270.683	-8.3	102	0.00
30 C	Chloroform	50.000	51.643	-3.3#	101	0.00
31 T	Cyclohexane	50.000	48.187	3.6	98	0.00
32 T	1,1,1-Trichloroethane	50.000	51.897	-3.8	100	0.00
33 S	1,2-Dichloroethane-d4	50.000	47.414	5.2	99	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	98	0.00
35 S	Dibromofluoromethane	50.000	47.868	4.3	99	0.00
36 T	1,1-Dichloropropene	50.000	48.644	2.7	100	0.00
37 T	Ethyl Acetate	50.000	52.947	-5.9	101	0.00
38 T	Carbon Tetrachloride	50.000	49.897	0.2	99	0.00
39 T	Methylcyclohexane	50.000	47.542	4.9	97	0.00
40 TM	Benzene	50.000	49.562	0.9	100	0.00
41 T	Methacrylonitrile	50.000	54.369	-8.7	103	0.00
42 TM	1,2-Dichloroethane	50.000	50.191	-0.4	100	0.00
43 T	Isopropyl Acetate	50.000	51.462	-2.9	100	0.00
44 TM	Trichloroethene	50.000	49.642	0.7	99	0.00
45 C	1,2-Dichloropropane	50.000	50.674	-1.3#	100	0.00
46 T	Dibromomethane	50.000	50.065	-0.1	98	0.00
47 T	Bromodichloromethane	50.000	51.486	-3.0	99	0.00
48 T	Methyl methacrylate	50.000	53.874	-7.7	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
 Data File : VX041934.D
 Acq On : 18 Jun 2024 19:34
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Jun 19 08:50:39 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 19 07:57:22 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)

49 T	1,4-Dioxane	1000.000	1077.740	-7.8	114	0.00
50 S	Toluene-d8	50.000	48.072	3.9	99	0.00
51 T	4-Methyl-2-Pentanone	250.000	269.979	-8.0	102	0.00
52 CM	Toluene	50.000	50.276	-0.6#	99	0.00
53 T	t-1,3-Dichloropropene	50.000	46.097	7.8	93	0.00
54 T	cis-1,3-Dichloropropene	50.000	50.288	-0.6	94	0.00
55 T	1,1,2-Trichloroethane	50.000	50.681	-1.4	97	0.00
56 T	Ethyl methacrylate	50.000	52.526	-5.1	97	0.00
57 T	1,3-Dichloropropane	50.000	51.172	-2.3	100	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	266.565	-6.6	97	0.00
59 T	2-Hexanone	250.000	274.921	-10.0	102	0.00
60 T	Dibromochloromethane	50.000	52.024	-4.0	97	0.00
61 T	1,2-Dibromoethane	50.000	51.696	-3.4	100	0.00
62 S	4-Bromofluorobenzene	50.000	47.785	4.4	94	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	98	0.00
64 T	Tetrachloroethene	50.000	49.995	0.0	103	0.00
65 PM	Chlorobenzene	50.000	49.763	0.5	99	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	50.429	-0.9	96	0.00
67 C	Ethyl Benzene	50.000	51.179	-2.4#	99	0.00
68 T	m/p-Xylenes	100.000	101.527	-1.5	99	0.00
69 T	o-Xylene	50.000	51.349	-2.7	98	0.00
70 T	Styrene	50.000	52.754	-5.5	98	0.00
71 P	Bromoform	50.000	48.054	3.9	94	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	96	0.00
73 T	Isopropylbenzene	50.000	52.506	-5.0	99	0.00
74 T	N-ethyl acetate	50.000	52.438	-4.9	97	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.688	-1.4	98	0.00
76 T	1,2,3-Trichloropropane	50.000	55.560	-11.1	110	0.00
77 T	Bromobenzene	50.000	50.225	-0.5	96	0.00
78 T	n-propylbenzene	50.000	52.328	-4.7	97	0.00
79 T	2-Chlorotoluene	50.000	51.188	-2.4	99	0.00
80 T	1,3,5-Trimethylbenzene	50.000	52.472	-4.9	99	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	46.179	7.6	85	0.00
82 T	4-Chlorotoluene	50.000	52.068	-4.1	97	0.00
83 T	tert-Butylbenzene	50.000	52.035	-4.1	98	0.00
84 T	1,2,4-Trimethylbenzene	50.000	53.084	-6.2	98	0.00
85 T	sec-Butylbenzene	50.000	52.908	-5.8	98	0.00
86 T	p-Isopropyltoluene	50.000	52.376	-4.8	96	0.00
87 T	1,3-Dichlorobenzene	50.000	50.974	-1.9	97	0.00
88 T	1,4-Dichlorobenzene	50.000	49.582	0.8	96	0.00
89 T	n-Butylbenzene	50.000	51.860	-3.7	95	0.00
90 T	Hexachloroethane	50.000	51.560	-3.1	97	0.00
91 T	1,2-Dichlorobenzene	50.000	51.293	-2.6	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	54.896	-9.8	99	0.00
93 T	1,2,4-Trichlorobenzene	50.000	50.168	-0.3	93	0.00
94 T	Hexachlorobutadiene	50.000	48.970	2.1	95	0.00
95 T	Naphthalene	50.000	48.637	2.7	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041934.D
Acq On : 18 Jun 2024 19:34
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Jun 19 08:50:39 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 19 07:57:22 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	50.901	-1.8	94	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6