


Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061824\
Data File : VX041910.D
Acq On : 18 Jun 2024 08:37
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061824W.M
Title : SW846 8260
Last Update : Wed Jun 19 07:57:22 2024



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.1	9806	PASS
75	95	30	60	46.1	28125	PASS
95	95	100	100	100.0	60947	PASS
96	95	5	9	6.9	4179	PASS
173	174	0.00	2	0.7	428	PASS
174	95	50	100	94.8	57781	PASS
175	174	5	9	7.6	4377	PASS
176	174	95	101	97.6	56376	PASS
177	176	5	9	6.5	3641	PASS