

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
 Data File : VX046739.D
 Acq On : 18 Jun 2025 13:58
 Operator : JC/MD
 Sample : Q2345-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
ClientSampleId :
 BPOW6-5-20250616

Quant Time: Jun 19 01:24:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	5.568	168	138024	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	229333	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204785	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	98514	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	92104	48.244	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	96.480%
35) Dibromofluoromethane	5.403	113	73789	48.631	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.260%
50) Toluene-d8	8.653	98	268168	49.192	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.380%
62) 4-Bromofluorobenzene	11.079	95	101037	49.720	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	99.440%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
 Data File : VX046739.D
 Acq On : 18 Jun 2025 13:58
 Operator : JC/MD
 Sample : Q2345-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BPOW6-5-20250616

Quant Time: Jun 19 01:24:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

