

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
 Data File : VX046743.D
 Acq On : 18 Jun 2025 15:24
 Operator : JC/MD
 Sample : Q2345-13
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TT189D2-20250617

Quant Time: Jun 19 01:25:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	217878	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	358276	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	311829	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	149381	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	142368	47.241	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	94.480%
35) Dibromofluoromethane	5.397	113	116087	48.972	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.940%
50) Toluene-d8	8.653	98	415275	48.761	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.520%
62) 4-Bromofluorobenzene	11.079	95	153648	48.398	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.800%
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.185	85	968	0.416	ug/l	90
9) 1,1,2-Trichlorotrifluo...	2.349	101	30874	12.379	ug/l	99
12) 1,1-Dichloroethene	2.337	96	1932	0.803	ug/l	96
27) cis-1,2-Dichloroethene	4.507	96	5156	1.704	ug/l	91
44) Trichloroethene	7.135	130	151205	58.588	ug/l	99
64) Tetrachloroethene	9.281	164	724	0.335	ug/l	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
Data File : VX046743.D
Acq On : 18 Jun 2025 15:24
Operator : JC/MD
Sample : Q2345-13
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TT189D2-20250617

Quant Time: Jun 19 01:25:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

