

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
 Data File : VX046745.D
 Acq On : 18 Jun 2025 16:07
 Operator : JC/MD
 Sample : Q2345-15
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW8-MW01D1-20250617

Quant Time: Jun 19 01:26:25 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	226176	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	367631	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	322256	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	153646	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.970	65	143488	45.866	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.740%
35) Dibromofluoromethane	5.403	113	119388	49.083	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.160%
50) Toluene-d8	8.653	98	430858	49.303	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.600%
62) 4-Bromofluorobenzene	11.079	95	157725	48.418	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	96.840%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
Data File : VX046745.D
Acq On : 18 Jun 2025 16:07
Operator : JC/MD
Sample : Q2345-15
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW8-MW01D1-20250617

Quant Time: Jun 19 01:26:25 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

