

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
 Data File : VX046749.D
 Acq On : 18 Jun 2025 17:34
 Operator : JC/MD
 Sample : Q2344-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-1

Quant Time: Jun 19 01:27:53 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	206985	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	342093	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	301676	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	143592	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	134324	46.918	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	93.840%		
35) Dibromofluoromethane	5.403	113	110764	48.937	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.880%		
50) Toluene-d8	8.653	98	402759	49.529	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	99.060%		
62) 4-Bromofluorobenzene	11.079	95	145810	48.102	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.200%		
Target Compounds						
16) Acetone	2.398	43	997	1.044	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
Data File : VX046749.D
Acq On : 18 Jun 2025 17:34
Operator : JC/MD
Sample : Q2344-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-1

Quant Time: Jun 19 01:27:53 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

