

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
 Data File : VX046752.D
 Acq On : 18 Jun 2025 18:37
 Operator : JC/MD
 Sample : Q2344-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2

Quant Time: Jun 19 01:28:54 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 18 03:09:16 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.568	168	213961	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	355243	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	307080	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	148475	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	137712	46.533	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.060%
35) Dibromofluoromethane	5.397	113	114074	48.534	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.060%
50) Toluene-d8	8.653	98	413878	49.012	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.020%
62) 4-Bromofluorobenzene	11.079	95	150781	47.901	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	95.800%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	4.513	96	1468	0.494	ug/l	82
65) Chlorobenzene	10.079	112	3095	0.457	ug/l #	86

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061825\
Data File : VX046752.D
Acq On : 18 Jun 2025 18:37
Operator : JC/MD
Sample : Q2344-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2

Quant Time: Jun 19 01:28:54 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061725W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 18 03:09:16 2025
Response via : Initial Calibration

