

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016919.D
 Acq On : 20 Jun 2020 09:17
 Operator : JC/SP
 Sample : L3066-06MSD
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD

Quant Time: Jun 22 01:59:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	367372	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	585880	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	549098	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	262140	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	179935	40.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	81.44%	
35) Dibromofluoromethane	5.47	113	163538	43.63	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.26%	
50) Toluene-d8	8.70	98	651377	43.44	ug/l	0.00
Spiked Amount	50.000		Recovery	=	86.88%	
62) 4-Bromofluorobenzene	11.12	95	244292	42.52	ug/l	0.00
Spiked Amount	50.000		Recovery	=	85.04%	

Target Compounds

						Qvalue
2) Dichlorodifluoromethane	1.18	85	212192	47.066	ug/l	98
3) Chloromethane	1.31	50	175465	38.383	ug/l	100
4) Vinyl Chloride	1.39	62	248760	48.281	ug/l	98
5) Bromomethane	1.62	94	81907	26.402	ug/l	100
6) Chloroethane	1.71	64	156628	49.359	ug/l	99
7) Trichlorofluoromethane	1.92	101	316571	43.920	ug/l	99
8) Diethyl Ether	2.17	74	132115	47.144	ug/l	96
9) 1,1,2-Trichlorotrifluoroet	2.36	101	160747	41.869	ug/l	98
10) Methyl Iodide	2.49	142	96062	26.283	ug/l	94
11) Tert butyl alcohol	3.01	59	191354	256.191	ug/l	99
12) 1,1-Dichloroethene	2.36	96	172959	44.056	ug/l	98
13) Acrolein	2.27	56	143110	316.018	ug/l	100
14) Allyl chloride	2.71	41	252617	48.646	ug/l	95
15) Acrylonitrile	3.12	53	468634	273.793	ug/l	98
16) Acetone	2.42	43	367724	229.266	ug/l	98
17) Carbon Disulfide	2.55	76	460705	45.409	ug/l	100
18) Methyl Acetate	2.75	43	197369	48.392	ug/l	99
19) Methyl tert-butyl Ether	3.17	73	552005	49.339	ug/l	99
20) Methylene Chloride	2.84	84	177766	46.217	ug/l	98
21) trans-1,2-Dichloroethene	3.14	96	174152	46.994	ug/l	98
22) Diisopropyl ether	3.82	45	528605	53.130	ug/l	96
23) Vinyl Acetate	3.78	43	1870513	227.996	ug/l	98
24) 1,1-Dichloroethane	3.67	63	306291	49.284	ug/l	99
25) 2-Butanone	4.64	43	707791	320.644	ug/l	97
26) 2,2-Dichloropropane	4.55	77	190157	32.658	ug/l	97
27) cis-1,2-Dichloroethene	4.56	96	225837	54.220	ug/l	98
28) Bromochloromethane	4.98	49	148312	57.498	ug/l	90
29) Tetrahydrofuran	5.09	42	453859	326.466	ug/l	95
30) Chloroform	5.18	83	350641	53.441	ug/l	98
31) Cyclohexane	5.55	56	291419	53.336	ug/l	98
32) 1,1,1-Trichloroethane	5.46	97	305235	49.152	ug/l	99
36) 1,1-Dichloropropene	5.77	75	261529	48.345	ug/l	98
37) Ethyl Acetate	4.79	43	248376	51.965	ug/l	98
38) Carbon Tetrachloride	5.75	117	269574	44.412	ug/l	100

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\WX061920\
 Data File : VX016919.D
 Acq On : 20 Jun 2020 09:17
 Operator : JC/SP
 Sample : L3066-06MSD
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD

Quant Time: Jun 22 01:59:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
39) Methylcyclohexane	7.44	83	271792	43.210	ug/l	99
40) Benzene	6.12	78	732160	45.106	ug/l	99
41) Methacrylonitrile	5.01	41	154823	59.666	ug/l	97
42) 1,2-Dichloroethane	6.16	62	238595	41.308	ug/l	99
43) Isopropyl Acetate	6.42	43	369605	44.282	ug/l	99
44) Trichloroethene	7.19	130	226330	45.529	ug/l	98
45) 1,2-Dichloropropane	7.49	63	205357	50.326	ug/l	97
46) Dibromomethane	7.64	93	134330	48.543	ug/l	97
47) Bromodichloromethane	7.88	83	279961	47.569	ug/l	99
48) Methyl methacrylate	7.74	41	218187	55.513	ug/l	98
49) 1,4-Dioxane	7.71	88	88539	1006.332	ug/l	97
51) 4-Methyl-2-Pentanone	8.62	43	1374932	273.879	ug/l	99
52) Toluene	8.77	92	520549	48.342	ug/l	100
53) t-1,3-Dichloropropene	9.02	75	314694	48.637	ug/l	99
54) cis-1,3-Dichloropropene	8.42	75	321861	46.499	ug/l	100
55) 1,1,2-Trichloroethane	9.20	97	207740	51.097	ug/l	97
56) Ethyl methacrylate	9.16	69	334515	56.836	ug/l	98
57) 1,3-Dichloropropane	9.35	76	356061	51.661	ug/l	100
59) 2-Hexanone	9.48	43	1023631	281.567	ug/l	99
60) Dibromochloromethane	9.57	129	240486	49.308	ug/l	100
61) 1,2-Dibromoethane	9.65	107	222345	50.394	ug/l	99
64) Tetrachloroethene	9.32	164	187566	36.888	ug/l	91
65) Chlorobenzene	10.12	112	574227	51.643	ug/l	100
66) 1,1,1,2-Tetrachloroethane	10.21	131	215635	50.039	ug/l	97
67) Ethyl Benzene	10.24	91	2222956	118.491	ug/l	100
68) m/p-Xylenes	10.35	106	1359582	186.235	ug/l	99
69) o-Xylene	10.68	106	373555	55.339	ug/l	97
70) Styrene	10.70	104	638404	55.661	ug/l	99
71) Bromoform	10.84	173	181260	50.029	ug/l #	100
73) Isopropylbenzene	11.01	105	4175799	247.411	ug/l	99
74) N-amyl acetate	10.88	43	363289	61.641	ug/l #	87
75) 1,1,2,2-Tetrachloroethane	11.26	83	307163	64.451	ug/l	96
76) 1,2,3-Trichloropropane	11.28	75	339859	70.988	ug/l	86
77) Bromobenzene	11.24	156	245681	53.374	ug/l	96
78) n-propylbenzene	11.35	91	7072339	375.465	ug/l	99
79) 2-Chlorotoluene	11.41	91	1018069	88.338	ug/l	77
80) 1,3,5-Trimethylbenzene	11.49	105	2920950	200.582	ug/l	99
81) trans-1,4-Dichloro-2-buten	11.06	75	90235	46.156	ug/l	99
82) 4-Chlorotoluene	11.49	91	922477	66.720	ug/l	87
83) tert-Butylbenzene	11.76	119	1057942	73.426	ug/l	98
84) 1,2,4-Trimethylbenzene	11.79	105	13811466	920.718	ug/l	80
85) sec-Butylbenzene	11.93	105	2051138	122.575	ug/l	84
86) p-Isopropyltoluene	12.05	119	1187904	77.479	ug/l	99
87) 1,3-Dichlorobenzene	12.01	146	408942	48.628	ug/l	99
88) 1,4-Dichlorobenzene	12.09	146	406523	47.799	ug/l	99
89) n-Butylbenzene	12.37	91	1246186	94.785	ug/l #	80
90) Hexachloroethane	12.58	117	267085	96.706	ug/l #	49
91) 1,2-Dichlorobenzene	12.38	146	396512	48.841	ug/l	98
92) 1,2-Dibromo-3-Chloropropan	12.98	75	84354	74.648	ug/l	84
93) 1,2,4-Trichlorobenzene	13.63	180	278617	51.108	ug/l	97


Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016919.D
Acq On : 20 Jun 2020 09:17
Operator : JC/SP
Sample : L3066-06MSD
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD

Quant Time: Jun 22 01:59:47 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
94) Hexachlorobutadiene	13.77	225	72916	29.641	ug/l	97
95) Naphthalene	13.82	128	2024158	126.471	ug/l	99
96) 1,2,3-Trichlorobenzene	14.00	180	281759	51.826	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.63 min Scan# 745

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion:168 Resp: 367372

Ion Ratio Lower Upper

168 100

99 51.5 42.2 63.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.63

150000



100000

50000

0

5.50 5.60 5.70 5.80

Time-->



#2

Dichlorodifluoromethane

Concen: 47.066 ug/l

RT: 1.18 min Scan# 16

Delta R.T. 0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 85 Resp: 212192

Ion Ratio Lower Upper

85 100

87 32.8 16.0 47.9



Abundance Ion 84.90 (84.60 to 85.60): VX0

Ion 86.90 (86.60 to 87.60): VX0

1.18

200000

150000


100000


50000

0

1.10 1.20 1.30

Time-->





#3

Chloromethane

Concen: 38.383 ug/l

RT: 1.31 min Scan# 37

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion: 50 Resp: 175465

Ion Ratio Lower Upper

50 100

52 32.7 26.1 39.1



Abundance Ion 49.90 (49.60 to 50.60): VX0

Ion 51.90 (51.60 to 52.60): VX0

1.31

200000



150000

100000

50000

0

Time--> 1.25 1.30 1.35 1.40



#4

Vinyl Chloride

Concen: 48.281 ug/l

RT: 1.39 min Scan# 50

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 62 Resp: 248760

Ion Ratio Lower Upper

62 100

64 32.6 25.0 37.4



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 63.90 (63.60 to 64.60): VX0

1.39

250000

200000


150000


100000

50000

0

Time--> 1.30 1.40 1.50





#5

Bromomethane

Concen: 26.402 ug/l

RT: 1.62 min Scan# 88

Delta R.T. 0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion: 94 Resp: 81907

Ion Ratio Lower Upper




94 100

96 93.0 74.4 111.6



Abundance Ion 93.90 (93.60 to 94.60): VX0

Ion 95.90 (95.60 to 96.60): VX0



#6

Chloroethane

Concen: 49.359 ug/l

RT: 1.71 min Scan# 102

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 64 Resp: 156628

Ion Ratio Lower Upper



64 100


66 31.7 25.7 38.5



Abundance Ion 63.90 (63.60 to 64.60): VX0

Ion 65.90 (65.60 to 66.60): VX0









#7

Trichlorofluoromethane
Concen: 43.920 ug/l
RT: 1.92 min Scan# 136
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD




Tot Ion:101 Resp: 316571
Ion Ratio Lower Upper
101 100
103 65.1 51.2 76.8




#8

Diethyl Ether
Concen: 47.144 ug/l
RT: 2.17 min Scan# 177
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 74 Resp: 132115
Ion Ratio Lower Upper
74 100
45 94.7 45.3 135.8





#9

1,1,2-Trichlorotrifluoroethane

Concen: 41.869 ug/l

RT: 2.36 min Scan# 209

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion:101 Resp: 160747

Ion Ratio Lower Upper

101 100

85 42.0 34.3 51.5




151 82.1 67.6 101.4



Abundance Ion 100.90 (100.60 to 101.60): V

Ion 84.90 (84.60 to 85.60): VX0

Ion 150.80 (150.50 to 151.50): V



#10

Methyl Iodide

Concen: 26.283 ug/l

RT: 2.49 min Scan# 230

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion:142 Resp: 96062

Ion Ratio Lower Upper

142 100

127 39.3 35.1 52.7



141 14.2 11.6 17.4




Abundance Ion 141.80 (141.50 to 142.50): V

Ion 126.80 (126.50 to 127.50): V

Ion 140.80 (140.50 to 141.50): V









#11

Tert butyl alcohol
 Concen: 256.191 ug/l
 RT: 3.01 min Scan# 316
 Delta R.T. 0.01 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD




Tot Ion: 59 Resp: 191354
 Ion Ratio Lower Upper
 59 100
 57 10.0 8.4 12.6




#12

1,1-Dichloroethene
 Concen: 44.056 ug/l
 RT: 2.36 min Scan# 208
 Delta R.T. -0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Tgt Ion: 96 Resp: 172959
 Ion Ratio Lower Upper
 96 100
 61 156.0 121.7 182.5
 98 64.7 51.1 76.7





#13

Acrolein

Concen: 316.018 ug/l

RT: 2.27 min Scan# 194

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 56 Resp: 143110

Ion Ratio Lower Upper

56 100

55 69.8 55.7 83.5



#14

Allyl chloride

Concen: 48.646 ug/l

RT: 2.71 min Scan# 266

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17




Tgt Ion: 41 Resp: 252617


Ion Ratio Lower Upper

41 100

39 58.2 50.7 76.1

76 36.5 30.9 46.3





#15

Acrylonitrile

Concen: 273.793 ug/l

RT: 3.12 min Scan# 333

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD





Tot Ion: 53 Resp: 468634

Ion Ratio Lower Upper

53 100

52 81.2 66.2 99.2

51 34.8 28.7 43.1



#16

Acetone

Concen: 229.266 ug/l

RT: 2.42 min Scan# 218

Delta R.T. 0.00 min

Lab File: VX016919.D




Acq: 20 Jun 2020 09:17


Tgt Ion: 43 Resp: 367724

Ion Ratio Lower Upper

43 100

58 34.4 26.6 39.8





#17

Carbon Disulfide

Concen: 45.409 ug/l

RT: 2.55 min Scan# 240

Delta R.T. 0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion: 76 Resp: 460705

Ion Ratio Lower Upper

76 100

78 9.3 7.5 11.3



Abundance Ion 75.90 (75.60 to 76.60): VX0

Ion 77.90 (77.60 to 78.60): VX0

2.55

250000

200000

150000



100000

50000

0

Time-->

2.50 2.60 2.70



#18

Methyl Acetate

Concen: 48.392 ug/l

RT: 2.75 min Scan# 272

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 43 Resp: 197369

Ion Ratio Lower Upper

43 100

74 26.8 21.8 32.8



Abundance Ion 43.00 (42.70 to 43.70): VX0

Ion 74.00 (73.70 to 74.70): VX0

2.75

120000

100000

80000

60000


40000


20000

0

Time-->

2.65 2.70 2.75 2.80 2.85









#19
Methyl tert-butyl Ether
Concen: 49.339 ug/l
RT: 3.17 min Scan# 341
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion: 73 Resp: 552005
Ion Ratio Lower Upper
73 100
57 21.8 16.9 25.3



#20
Methylene Chloride
Concen: 46.217 ug/l
RT: 2.84 min Scan# 287
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 84 Resp: 177766
Ion Ratio Lower Upper
84 100
49 109.8 86.6 129.8
51 34.8 27.1 40.7
86 64.1 53.1 79.7





#21

trans-1,2-Dichloroethene

Concen: 46.994 ug/l

RT: 3.14 min Scan# 337

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion: 96 Resp: 174152

Ion Ratio Lower Upper

96 100

61 134.3 105.4 158.0




98 64.9 51.3 76.9



Abundance Ion 95.90 (95.60 to 96.60): VX0

Ion 60.90 (60.60 to 61.60): VX0

Ion 97.90 (97.60 to 98.60): VX0



#22

Diisopropyl ether

Concen: 53.130 ug/l

RT: 3.82 min Scan# 449

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Tgt Ion: 45 Resp: 528605


Ion Ratio Lower Upper

45 100

43 66.6 57.0 85.4

87 30.6 25.2 37.8

59 12.4 9.9 14.9






Abundance Ion 45.00 (44.70 to 45.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 87.00 (86.70 to 87.70): VX0

Ion 59.00 (58.70 to 59.70): VX0





#23

Vinyl Acetate

Concen: 227.996 ug/l

RT: 3.78 min Scan# 442

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 43 Resp: 1870513

Ion Ratio Lower Upper

43 100

86 11.7 9.9 14.9



#24

1,1-Dichloroethane

Concen: 49.284 ug/l

RT: 3.67 min Scan# 424

Delta R.T. 0.01 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17




Tgt Ion: 63 Resp: 306291


Ion Ratio Lower Upper

63 100

98 7.1 3.6 10.9

100 5.1 2.4 7.1









#25
2-Butanone
Concen: 320.644 ug/l
RT: 4.64 min Scan# 582
Delta R.T. 0.01 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampled :
MW-2MSD


Tot Ion: 43 Resp: 707791
Ion Ratio Lower Upper
43 100
72 27.8 23.5 35.3



#26
2,2-Dichloropropane
Concen: 32.658 ug/l
RT: 4.55 min Scan# 568
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 77 Resp: 190157
Ion Ratio Lower Upper
77 100
97 24.7 11.6 34.7









#27

cis-1,2-Dichloroethene
 Concen: 54.220 ug/l
 RT: 4.56 min Scan# 570
 Delta R.T. -0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD




Tot Ion	Ratio	Lower	Upper
96	100		
61	133.3	0.0	273.8
98	64.7	0.0	129.0




#28

Bromochloromethane
 Concen: 57.498 ug/l
 RT: 4.98 min Scan# 639
 Delta R.T. 0.01 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Tgt Ion	Ratio	Lower	Upper
49	100		
129	2.2	0.0	5.0
130	85.6	76.8	115.2





#29

Tetrahydrofuran

Concen: 326.466 ug/l

RT: 5.09 min Scan# 657

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion: 42 Resp: 453859

Ion Ratio Lower Upper

42 100

72 47.7 41.1 61.7




71 44.6 38.4 57.6



Abundance Ion 42.00 (41.70 to 42.70): VX0

Ion 72.00 (71.70 to 72.70): VX0

Ion 71.00 (70.70 to 71.70): VX0



#30

Chloroform

Concen: 53.441 ug/l

RT: 5.18 min Scan# 671

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 83 Resp: 350641

Ion Ratio Lower Upper



83 100


85 64.8 53.4 80.0



Abundance Ion 82.90 (82.60 to 83.60): VX0

Ion 84.90 (84.60 to 85.60): VX0









#31
Cyclohexane
Concen: 53.336 ug/l
RT: 5.55 min Scan# 732
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion: 56 Resp: 291419
Ion Ratio Lower Upper
56 100
69 32.0 26.1 39.1
84 88.8 73.0 109.6



#32
1,1,1-Trichloroethane
Concen: 49.152 ug/l
RT: 5.46 min Scan# 718
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 97 Resp: 305235
Ion Ratio Lower Upper
97 100
99 63.8 51.3 76.9
61 40.4 31.3 46.9





#33

1,2-Dichloroethane-d4

Concen: 40.725 ug/l

RT: 6.03 min Scan# 811

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 65 Resp: 179935

Ion Ratio Lower Upper

65 100

67 52.9 0.0 102.6



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.83 min Scan# 942

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17




Tgt Ion: 114 Resp: 585880


Ion Ratio Lower Upper

114 100

63 19.7 0.0 37.0

88 16.2 0.0 32.0





#35

Dibromofluoromethane

Concen: 43.635 ug/l

RT: 5.47 min Scan# 719

Delta R.T. 0.01 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion:113 Resp: 163538

Ion Ratio Lower Upper

113 100

111 103.0 81.0 121.4




192 23.0 18.5 27.7



Abundance Ion 112.80 (112.50 to 113.50): V

Ion 110.80 (110.50 to 111.50): V

Ion 191.70 (191.40 to 192.40): V



#36

1,1-Dichloropropene

Concen: 48.345 ug/l

RT: 5.77 min Scan# 768

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion: 75 Resp: 261529

Ion Ratio Lower Upper

75 100

110 36.8 18.9 56.9



77 31.1 25.6 38.4




Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 109.90 (109.60 to 110.60): V

Ion 76.90 (76.60 to 77.60): VX0









#37
Ethyl Acetate
Concen: 51.965 ug/l
RT: 4.79 min Scan# 608
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampled :
MW-2MSD


Tot Ion: 43 Resp: 248376
Ion Ratio Lower Upper
43 100
61 14.8 12.2 18.2
70 12.2 10.5 15.7



#38
Carbon Tetrachloride
Concen: 44.412 ug/l
RT: 5.75 min Scan# 765
Delta R.T. -0.01 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 117 Resp: 269574
Ion Ratio Lower Upper
117 100
119 96.5 77.4 116.2
121 31.3 24.7 37.1









#39
Methvlcvclohexane
Concen: 43.210 ug/l
RT: 7.44 min Scan# 1042
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion: 83 Resp: 271792
Ion Ratio Lower Upper
83 100
55 75.9 61.7 92.5
98 49.6 39.0 58.6



#40
Benzene
Concen: 45.106 ug/l
RT: 6.12 min Scan# 825
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 78 Resp: 732160
Ion Ratio Lower Upper
78 100
77 23.8 18.6 27.8





#41

Methacrylonitrile

Concen: 59.666 ug/l

RT: 5.01 min Scan# 643

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD

Tot Ion: 41 Resp: 154823


Ion Ratio Lower Upper

41 100

39 53.1 42.6 64.0

67 76.9 64.2 96.2

52 29.3 25.6 38.4






Abundance Ion 41.00 (40.70 to 41.70): VX0

Ion 39.00 (38.70 to 39.70): VX0

Ion 67.00 (66.70 to 67.70): VX0

Ion 52.00 (51.70 to 52.70): VX0



#42

1,2-Dichloroethane

Concen: 41.308 ug/l

RT: 6.16 min Scan# 832

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 62 Resp: 238595

Ion Ratio Lower Upper



62 100


98 10.9 0.0 20.8



Abundance Ion 61.90 (61.60 to 62.60): VX0

Ion 97.90 (97.60 to 98.60): VX0






#43




Isopropyl Acetate
 Concen: 44.282 ug/l
 RT: 6.42 min Scan# 874
 Delta R.T. 0.01 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD

Tot Ion	Ratio	Lower	Upper
43	100		
61	22.5	17.9	26.9
87	15.1	12.4	18.6




Abundance Ion 43.00 (42.70 to 43.70): VX0
 Ion 61.00 (60.70 to 61.70): VX0
 Ion 87.00 (86.70 to 87.70): VX0





#44


Trichloroethene
 Concen: 45.529 ug/l
 RT: 7.19 min Scan# 1001
 Delta R.T. -0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Tgt Ion	Ratio	Lower	Upper
130	100		
95	96.9	0.0	190.8



Abundance Ion 129.80 (129.50 to 130.50): V
 Ion 94.90 (94.60 to 95.60): VX0





#45

1,2-Dichloropropane

Concen: 50.326 ug/l

RT: 7.49 min Scan# 1050

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 63 Resp: 205357

Ion Ratio Lower Upper

63 100

65 30.5 25.8 38.8



#46

Dibromomethane

Concen: 48.543 ug/l

RT: 7.64 min Scan# 1075

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17




Tgt Ion: 93 Resp: 134330


Ion Ratio Lower Upper

93 100

95 83.7 65.7 98.5

174 112.7 92.8 139.2





#47

Bromodichloromethane

Concen: 47.569 ug/l

RT: 7.88 min Scan# 1114

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion: 83 Resp: 279961

Ion Ratio Lower Upper




83 100

85 64.6 50.9 76.3

127 9.1 7.4 11.0



Abundance Ion 82.90 (82.60 to 83.60): VX0
Ion 84.90 (84.60 to 85.60): VX0
Ion 126.80 (126.50 to 127.50): V



#48

Methyl methacrylate

Concen: 55.513 ug/l

RT: 7.74 min Scan# 1092

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion: 41 Resp: 218187

Ion Ratio Lower Upper



41 100


69 90.5 73.3 109.9

39 50.5 42.4 63.6



Abundance Ion 41.00 (40.70 to 41.70): VX0
Ion 69.00 (68.70 to 69.70): VX0
Ion 39.00 (38.70 to 39.70): VX0





#49

1,4-Dioxane

Concen: 1006.332 ug/l

RT: 7.71 min Scan# 1087

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion: 88 Resp: 88539

Ion Ratio Lower Upper

88 100

43 29.1 22.7 34.1




58 68.9 53.1 79.7



Abundance Ion 88.00 (87.70 to 88.70): VX0

Ion 43.00 (42.70 to 43.70): VX0

Ion 58.00 (57.70 to 58.70): VX0



#50

Toluene-d8

Concen: 43.440 ug/l

RT: 8.70 min Scan# 1248

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 98 Resp: 651377

Ion Ratio Lower Upper



98 100


100 67.6 54.0 81.0



Abundance Ion 98.00 (97.70 to 98.70): VX0

Ion 100.00 (99.70 to 100.70): V





#51

4-Methyl-2-Pentanone

Concen: 273.879 ug/l

RT: 8.62 min Scan# 1236

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 43 Resp: 1374932

Ion Ratio Lower Upper

43 100

58 42.3 34.3 51.5



#52

Toluene

Concen: 48.342 ug/l

RT: 8.77 min Scan# 1260

Delta R.T. 0.01 min

Lab File: VX016919.D




Acq: 20 Jun 2020 09:17


Tgt Ion: 92 Resp: 520549

Ion Ratio Lower Upper

92 100

91 168.3 134.2 201.2









#53
t-1,3-Dichloropropene
Concen: 48.637 ug/l
RT: 9.02 min Scan# 1302
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion: 75 Resp: 314694
Ion Ratio Lower Upper
75 100
77 31.2 25.3 37.9



#54
cis-1,3-Dichloropropene
Concen: 46.499 ug/l
RT: 8.42 min Scan# 1202
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 75 Resp: 321861
Ion Ratio Lower Upper
75 100
77 30.9 24.9 37.3
39 42.1 33.4 50.0









#55
 1,1,2-Trichloroethane
 Concen: 51.097 ug/l
 RT: 9.20 min Scan# 1330
 Delta R.T. -0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD


Tot Ion:	97	Resp:	207740
Ion	Ratio	Lower	Upper
97	100		
83	86.2	65.5	98.3
85	54.3	42.6	63.8
99	63.8	49.4	74.0



#56
 Ethyl methacrylate
 Concen: 56.836 ug/l
 RT: 9.16 min Scan# 1324
 Delta R.T. 0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Tgt Ion:	69	Resp:	334515
Ion	Ratio	Lower	Upper
69	100		
41	62.5	47.9	71.9
39	30.1	24.2	36.4





#57

1,3-Dichloropropane

Concen: 51.661 ug/l

RT: 9.35 min Scan# 1356

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 76 Resp: 356061

Ion Ratio Lower Upper

76 100

78 32.1 25.8 38.6



#59

2-Hexanone

Concen: 281.567 ug/l

RT: 9.48 min Scan# 1376

Delta R.T. 0.01 min

Lab File: VX016919.D




Acq: 20 Jun 2020 09:17


Tgt Ion: 43 Resp: 1023631

Ion Ratio Lower Upper

43 100

58 59.2 29.8 89.4





#60

Dibromochloromethane

Concen: 49.308 ug/l

RT: 9.57 min Scan# 1391

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion:129 Resp: 240486

Ion Ratio Lower Upper

129 100

127 77.4 38.5 115.5



Abundance Ion 128.80 (128.50 to 129.50): V

Ion 126.80 (126.50 to 127.50): V

9.57



150000

100000

50000

0

Time-->



#61

1,2-Dibromoethane

Concen: 50.394 ug/l

RT: 9.65 min Scan# 1405

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion:107 Resp: 222345

Ion Ratio Lower Upper

107 100

109 94.9 75.4 113.0



Abundance Ion 106.90 (106.60 to 107.60): V

Ion 108.80 (108.50 to 109.50): V

9.65


150000


100000

50000

0

Time-->





#62

4-Bromofluorobenzene

Concen: 42.522 ug/l

RT: 11.12 min Scan# 1646

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion: 95 Resp: 244292

Ion Ratio Lower Upper

95 100

174 83.7 0.0 174.2




176 81.6 0.0 170.0



Abundance Ion 94.90 (94.60 to 95.60): VX0

Ion 173.80 (173.50 to 174.50): V

Ion 175.80 (175.50 to 176.50): V



#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 10.10 min Scan# 1478

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion: 117 Resp: 549098

Ion Ratio Lower Upper

117 100

82 53.8 43.4 65.0



119 31.8 25.7 38.5




Abundance Ion 116.90 (116.60 to 117.60): V

Ion 82.00 (81.70 to 82.70): VX0

Ion 118.90 (118.60 to 119.60): V





#64

Tetrachloroethene

Concen: 36.888 ug/l

RT: 9.32 min Scan# 1350

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD

Tot Ion:164 Resp: 187566


Ion Ratio Lower Upper

164 100

166 138.0 102.6 153.8

129 98.6 72.2 108.2

131 97.0 70.8 106.2






Abundance Ion 163.80 (163.50 to 164.50): V

Ion 165.80 (165.50 to 166.50): V

Ion 128.80 (128.50 to 129.50): V

Ion 130.80 (130.50 to 131.50): V



#65

Chlorobenzene

Concen: 51.643 ug/l

RT: 10.12 min Scan# 1482

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion:112 Resp: 574227

Ion Ratio Lower Upper



112 100


114 32.0 25.5 38.3



Abundance Ion 111.90 (111.60 to 112.60): V

Ion 113.90 (113.60 to 114.60): V





#66

1,1,1,2-Tetrachloroethane

Concen: 50.039 ug/l

RT: 10.21 min Scan# 1496

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion:131 Resp: 215635

Ion Ratio Lower Upper

131 100

133 97.2 47.7 143.1




119 64.7 30.8 92.4



Abundance Ion 130.80 (130.50 to 131.50): V

Ion 132.80 (132.50 to 133.50): V

Ion 118.80 (118.50 to 119.50): V



#67

Ethyl Benzene

Concen: 118.491 ug/l

RT: 10.24 min Scan# 1501

Delta R.T. -0.00 min

Lab File: VX016919.D


Acq: 20 Jun 2020 09:17

Tgt Ion: 91 Resp: 2222956

Ion Ratio Lower Upper



91 100


106 31.5 25.1 37.7



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 106.00 (105.70 to 106.70): V









#68
m/p-Xylenes
Concen: 186.235 ug/l
RT: 10.35 min Scan# 1519
Delta R.T. 0.01 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion:106 Resp: 1359582
Ion Ratio Lower Upper
106 100
91 200.9 161.5 242.3



#69
o-Xylene
Concen: 55.339 ug/l
RT: 10.68 min Scan# 1574
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion:106 Resp: 373555
Ion Ratio Lower Upper
106 100
91 211.4 107.9 323.7









#70
Stvrene
Concen: 55.661 ug/l
RT: 10.70 min Scan# 1576
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion:104 Resp: 638404
Ion Ratio Lower Upper
104 100
78 48.5 39.6 59.4
103 53.9 43.3 64.9



#71
Bromoform
Concen: 50.029 ug/l
RT: 10.84 min Scan# 1600
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion:173 Resp: 181260
Ion Ratio Lower Upper
173 100
175 48.7 24.3 73.0
254 0.0 0.1 0.1#









#72

1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.06 min Scan# 1800
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD




Tot Ion	Ratio	Lower	Upper
152	100		
115	88.3	40.9	122.8
150	168.2	0.0	344.2




#73

Isopropylbenzene
Concen: 247.411 ug/l
RT: 11.01 min Scan# 1627
Delta R.T. 0.01 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion	Ratio	Lower	Upper
105	100		
120	26.8	13.2	39.6





#74

N-amvl acetate

Concen: 61.641 ug/l

RT: 10.88 min Scan# 1606

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17





Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD

Tot Ion:	43	Resp:	363289
Ion	Ratio	Lower	Upper
43	100		
70	50.8	39.9	59.9
55	49.1	21.8	32.6
61	25.6	22.2	33.4



#75

1,1,2,2-Tetrachloroethane

Concen: 64.451 ug/l




RT: 11.26 min Scan# 1668


Delta R.T. 0.01 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Tgt Ion:	83	Resp:	307163
Ion	Ratio	Lower	Upper
83	100		
131	10.3	5.7	17.1
85	61.6	32.6	97.8





#76

1,2,3-Trichloropropane

Concen: 70.988 ug/l

RT: 11.28 min Scan# 1672

Delta R.T. 0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion: 75 Resp: 339859

Ion Ratio Lower Upper

75 100

77 32.3 20.5 61.5



Abundance Ion 74.90 (74.60 to 75.60): VX0

Ion 77.00 (76.70 to 77.70): VX0

300000



200000

100000

0

11.25 11.30

Time-->



#77

Bromobenzene

Concen: 53.374 ug/l

RT: 11.24 min Scan# 1665

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion: 156 Resp: 245681

Ion Ratio Lower Upper

156 100

77 153.2 73.4 220.2

158 96.2 48.8 146.3



Abundance Ion 155.80 (155.50 to 156.50): V

Ion 77.00 (76.70 to 77.70): VX0

Ion 157.80 (157.50 to 158.50): V

400000

300000


200000


100000

0

11.20 11.30

Time-->





#78

n-propylbenzene

Concen: 375.465 ug/l

RT: 11.35 min Scan# 1683

Delta R.T. 0.01 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion: 91 Resp: 7072339

Ion Ratio Lower Upper

91 100

120 24.3 11.9 35.5



#79

2-Chlorotoluene

Concen: 88.338 ug/l

RT: 11.41 min Scan# 1693

Delta R.T. 0.01 min

Lab File: VX016919.D




Acq: 20 Jun 2020 09:17


Tgt Ion: 91 Resp: 1018069

Ion Ratio Lower Upper

91 100

126 21.0 17.2 51.5









#80

1,3,5-Trimethylbenzene
Concen: 200.582 ug/l
RT: 11.49 min Scan# 1707
Delta R.T. 0.01 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Instrument :
MSVOA_X
ClientSampled :
MW-2MSD




Tot Ion:105 Resp: 2920950
Ion Ratio Lower Upper
105 100
120 49.8 24.6 74.0




#81

trans-1,4-Dichloro-2-butene
Concen: 46.156 ug/l
RT: 11.06 min Scan# 1636
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion: 75 Resp: 90235
Ion Ratio Lower Upper
75 100
53 87.0 69.8 104.8
89 46.8 38.3 57.5





#82

4-Chlorotoluene

Concen: 66.720 ug/l

RT: 11.49 min Scan# 1707

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :


MW-2MSD

Tot Ion: 91 Resp: 922477

Ion Ratio Lower Upper




91 100

126 23.0 15.0 45.0



Abundance Ion 91.00 (90.70 to 91.70): VX0

Ion 125.90 (125.60 to 126.60): V



#83

tert-Butylbenzene

Concen: 73.426 ug/l

RT: 11.76 min Scan# 1750

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion: 119 Resp: 1057942

Ion Ratio Lower Upper

119 100

91 58.7 28.5 85.5



134 23.8 11.7 35.1




Abundance Ion 119.00 (118.70 to 119.70): V

Ion 91.00 (90.70 to 91.70): VX0

Ion 134.00 (133.70 to 134.70): V









#84
 1,2,4-Trimethylbenzene
 Concen: 920.718 ug/l
 RT: 11.79 min Scan# 1756
 Delta R.T. -0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD


Tot Ion:105 Resp:13811466
 Ion Ratio Lower Upper
 105 100
 120 57.9 22.3 66.9



#85
 sec-Butylbenzene
 Concen: 122.575 ug/l
 RT: 11.93 min Scan# 1779
 Delta R.T. -0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Tgt Ion:105 Resp: 2051138
 Ion Ratio Lower Upper
 105 100
 134 28.2 10.4 31.2









#86
p-Isopropyltoluene
Concen: 77.479 ug/l
RT: 12.05 min Scan# 1798
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17




Instrument :
MSVOA_X
ClientSampleId :
MW-2MSD


Tot Ion:119 Resp: 1187904
Ion Ratio Lower Upper
119 100
134 26.1 12.9 38.7
91 22.3 11.6 34.8



#87
1,3-Dichlorobenzene
Concen: 48.628 ug/l
RT: 12.01 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX016919.D
Acq: 20 Jun 2020 09:17

Tgt Ion:146 Resp: 408942
Ion Ratio Lower Upper
146 100
111 38.7 19.4 58.2
148 64.6 31.8 95.3









#88
 1,4-Dichlorobenzene
 Concen: 47.799 ug/l
 RT: 12.09 min Scan# 1804
 Delta R.T. 0.01 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17




Instrument :
 MSVOA_X
 ClientSampleId :
 MW-2MSD


Tot Ion:146 Resp: 406523
 Ion Ratio Lower Upper
 146 100
 111 38.0 19.0 57.0
 148 64.1 31.6 95.0



#89
 n-Butylbenzene
 Concen: 94.785 ug/l
 RT: 12.37 min Scan# 1851
 Delta R.T. 0.00 min
 Lab File: VX016919.D
 Acq: 20 Jun 2020 09:17

Tgt Ion: 91 Resp: 1246186
 Ion Ratio Lower Upper
 91 100
 92 49.7 27.4 82.2
 134 51.1 13.9 41.6#





#90

Hexachloroethane

Concen: 96.706 ug/l

RT: 12.58 min Scan# 1885

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :





MW-2MSD

Tot Ion:117 Resp: 267085

Ion Ratio Lower Upper

117 100

201 42.1 44.9 134.5#



#91

1,2-Dichlorobenzene

Concen: 48.841 ug/l

RT: 12.38 min Scan# 1852

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17




Tgt Ion:146 Resp: 396512


Ion Ratio Lower Upper

146 100

111 41.8 20.9 62.7

148 63.5 32.9 98.6





#92

1,2-Dibromo-3-Chloropropane

Concen: 74.648 ug/l

RT: 12.98 min Scan# 1951

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion: 75 Resp: 84354

Ion Ratio Lower Upper

75 100

155 77.7 46.4 139.1




157 98.6 58.6 175.8



Abundance Ion 74.90 (74.60 to 75.60): VX016919.D (-1902) (-)

Ion 154.80 (154.50 to 155.50): VX016919.D (-1902) (-)

Ion 156.80 (156.50 to 157.50): VX016919.D (-1902) (-)



#93

1,2,4-Trichlorobenzene

Concen: 51.108 ug/l

RT: 13.63 min Scan# 2057

Delta R.T. 0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion: 180 Resp: 278617

Ion Ratio Lower Upper

180 100

182 96.4 47.7 143.1



145 35.9 15.0 45.1




Abundance Ion 179.80 (179.50 to 180.50): VX016919.D (-2008) (-)

Ion 181.80 (181.50 to 182.50): VX016919.D (-2008) (-)

Ion 144.80 (144.50 to 145.50): VX016919.D (-2008) (-)





#94

Hexachlorobutadiene

Concen: 29.641 ug/l

RT: 13.77 min Scan# 2080

Delta R.T. 0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD


Tot Ion:225 Resp: 72916

Ion Ratio Lower Upper




225 100

223 67.0 32.1 96.5

227 66.2 32.3 96.8



Abundance Ion 224.70 (224.40 to 225.40): V
Ion 222.70 (222.40 to 223.40): V
Ion 226.70 (226.40 to 227.40): V



#95

Naphthalene

Concen: 126.471 ug/l

RT: 13.82 min Scan# 2088

Delta R.T. -0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17


Tgt Ion:128 Resp: 2024158

Ion Ratio Lower Upper



128 100


127 12.9 10.4 15.6

129 11.1 8.6 13.0



Abundance Ion 128.00 (127.70 to 128.70): V
Ion 127.00 (126.70 to 127.70): V
Ion 129.00 (128.70 to 129.70): V





#96

1,2,3-Trichlorobenzene

Concen: 51.826 ug/l

RT: 14.00 min Scan# 2118

Delta R.T. 0.00 min

Lab File: VX016919.D

Acq: 20 Jun 2020 09:17

Instrument :

MSVOA_X

ClientSampleId :

MW-2MSD



Tot Ion:180 Resp: 281759

Ion Ratio Lower Upper

180 100

182 94.0 47.4 142.3

145 33.6 16.0 48.0



Abundance Ion 179.80 (179.50 to 180.50): V

Ion 181.80 (181.50 to 182.50): V

Ion 144.90 (144.60 to 145.60): V

