

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016877.D
 Acq On : 19 Jun 2020 14:21
 Operator : JC/SP
 Sample : L3042-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-200616

Quant Time: Jun 22 02:42:33 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	266619	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	424970	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	370022	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	191924	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	132002	41.17	ug/l	0.00
Spiked Amount	50.000		Recovery	=	82.34%	
35) Dibromofluoromethane	5.47	113	114016	41.94	ug/l	0.00
Spiked Amount	50.000		Recovery	=	83.88%	
50) Toluene-d8	8.70	98	541787	49.81	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.62%	
62) 4-Bromofluorobenzene	11.12	95	182933	43.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.80%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016877.D
Acq On : 19 Jun 2020 14:21
Operator : JC/SP
Sample : L3042-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-200616

Quant Time: Jun 22 02:42:33 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

