

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016881.D
 Acq On : 19 Jun 2020 15:52
 Operator : JC/SP
 Sample : L3032-03DL 20X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF015MW100-200616DL

Quant Time: Jun 22 03:34:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	220556	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	352109	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	359549	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	193186	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.03	65	130825	49.32	ug/l	0.00
Spiked Amount			Recovery	=	98.64%	
35) Dibromofluoromethane	5.46	113	111113	49.33	ug/l	0.00
Spiked Amount			Recovery	=	98.66%	
50) Toluene-d8	8.70	98	489475	54.32	ug/l	0.00
Spiked Amount			Recovery	=	108.64%	
62) 4-Bromofluorobenzene	11.12	95	182979	52.99	ug/l	0.00
Spiked Amount			Recovery	=	105.98%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
4) Vinyl Chloride	1.39	62	272345	88.045	ug/l	98
12) 1,1-Dichloroethene	2.36	96	25990	11.027	ug/l	99
21) trans-1,2-Dichloroethene	3.14	96	2250	1.011	ug/l #	91
27) cis-1,2-Dichloroethene	4.57	96	88381	35.343	ug/l	91
31) Cyclohexane	5.56	56	1200	0.366	ug/l #	79
39) Methylcyclohexane	7.45	83	1205	0.319	ug/l #	72
40) Benzene	6.11	78	8912	0.914	ug/l	99
44) Trichloroethene	7.19	130	117505	39.331	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016881.D
Acq On : 19 Jun 2020 15:52
Operator : JC/SP
Sample : L3032-03DL 20X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF015MW100-200616DL

Quant Time: Jun 22 03:34:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

