

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016884.D
 Acq On : 19 Jun 2020 17:00
 Operator : JC/SP
 Sample : L3032-02DL 10X
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LF015MW047-200616DL

Quant Time: Jun 22 03:50:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	214474	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	349658	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	350590	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	185479	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	130964	50.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.54%	
35) Dibromofluoromethane	5.47	113	107524	48.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.14%	
50) Toluene-d8	8.70	98	495651	55.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.78%	
62) 4-Bromofluorobenzene	11.12	95	208205	60.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	121.44%	
Target Compounds						
11) Tert butyl alcohol	3.01	59	6081	13.945	ug/l	95
22) Diisopropyl ether	3.82	45	6334	1.090	ug/l #	85
40) Benzene	6.12	78	10838	1.119	ug/l	91
65) Chlorobenzene	10.12	112	951080	133.966	ug/l	99
87) 1,3-Dichlorobenzene	12.01	146	10289	1.729	ug/l	97
88) 1,4-Dichlorobenzene	12.09	146	76490	12.711	ug/l	97
91) 1,2-Dichlorobenzene	12.38	146	297301	51.756	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016884.D
Acq On : 19 Jun 2020 17:00
Operator : JC/SP
Sample : L3032-02DL 10X
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LF015MW047-200616DL

Quant Time: Jun 22 03:50:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

