

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX061920\
 Data File : VX016886.D
 Acq On : 19 Jun 2020 17:45
 Operator : JC/SP
 Sample : L3020-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-01_061520

Quant Time: Jul 01 05:15:01 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	218801	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	348826	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	353140	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	217347	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	140565	53.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.84%	
35) Dibromofluoromethane	5.46	113	109618	49.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.24%	
50) Toluene-d8	8.70	98	427085	47.84	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.68%	
62) 4-Bromofluorobenzene	11.12	95	186914	54.64	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.28%	
Target Compounds						
16) Acetone	2.42	43	9720	10.175	ug/l	95
18) Methyl Acetate	2.75	43	1898	0.781	ug/l	97
20) Methylene Chloride	2.84	84	2066	0.902	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX061920\
Data File : VX016886.D
Acq On : 19 Jun 2020 17:45
Operator : JC/SP
Sample : L3020-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-01_061520

Quant Time: Jul 01 05:15:01 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

