

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016888.D
 Acq On : 19 Jun 2020 18:31
 Operator : JC/SP
 Sample : L3020-14
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02_061620

Quant Time: Jun 22 05:10:16 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	220298	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	359732	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	397217	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	191627	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	131996	49.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.64%	
35) Dibromofluoromethane	5.47	113	124729	54.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	108.40%	
50) Toluene-d8	8.70	98	418303	45.43	ug/l	0.00
Spiked Amount	50.000		Recovery	=	90.86%	
62) 4-Bromofluorobenzene	11.12	95	175546	49.76	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.52%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016888.D
Acq On : 19 Jun 2020 18:31
Operator : JC/SP
Sample : L3020-14
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02_061620

Quant Time: Jun 22 05:10:16 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

