

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016892.D
 Acq On : 19 Jun 2020 20:02
 Operator : JC/SP
 Sample : L3048-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GROUNDWATER

Quant Time: Jun 22 05:58:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	237616	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	387603	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	397628	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	211521	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	140443	49.14	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.28%	
35) Dibromofluoromethane	5.47	113	120034	48.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.82%	
50) Toluene-d8	8.70	98	473162	47.70	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.40%	
62) 4-Bromofluorobenzene	11.12	95	202652	53.32	ug/l	0.00
Spiked Amount	50.000		Recovery	=	106.64%	

Target Compounds

						Qvalue
16) Acetone	2.42	43	11263	10.857	ug/l	89
20) Methylene Chloride	2.84	84	1857	0.746	ug/l	95
30) Chloroform	5.18	83	10079	2.375	ug/l	98
44) Trichloroethene	7.19	130	2080	0.632	ug/l	87
51) 4-Methyl-2-Pentanone	8.62	43	2360	0.711	ug/l	93
67) Ethyl Benzene	10.24	91	4675	0.344	ug/l	99
84) 1,2,4-Trimethylbenzene	11.79	105	4693	0.388	ug/l	99
95) Naphthalene	13.82	128	4310	0.334	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016892.D
Acq On : 19 Jun 2020 20:02
Operator : JC/SP
Sample : L3048-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GROUNDWATER

Quant Time: Jun 22 05:58:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

