

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_X\DATA\VX061920\
 Data File : VX016894.D
 Acq On : 19 Jun 2020 20:47
 Operator : JC/SP
 Sample : L3066-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-5DL

Manual Integrations
 APPROVED

MMDadoda
 6/22/2020 12:23:19 PM

Quant Time: Jul 03 06:23:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	263686	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	394239	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	416662	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	241617	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	145888	46.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.00%	
35) Dibromofluoromethane	5.47	113	139330	55.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.50%	
50) Toluene-d8	8.70	98	480572	47.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.26%	
62) 4-Bromofluorobenzene	11.12	95	229879	59.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	118.92%	
Target Compounds						
16) Acetone	2.42	43	4030	3.501	ug/l #	86
67) Ethyl Benzene	10.24	91	73098	5.135	ug/l	98
68) m/p-Xylenes	10.34	106	6574	1.187	ug/l	94
69) o-Xylene	10.68	106	6484	1.266	ug/l	97
73) Isopropylbenzene	11.00	105	161846	10.404	ug/l	100
78) n-propylbenzene	11.35	91	290781	16.749	ug/l	99
80) 1,3,5-Trimethylbenzene	11.49	105	60141	4.481	ug/l	99
83) tert-Butylbenzene	11.76	119	18805	1.416	ug/l	90
84) 1,2,4-Trimethylbenzene	11.79	105	1120822	81.064	ug/l	97
85) sec-Butylbenzene	11.93	105	65215	4.228	ug/l	78
86) p-Isopropyltoluene	12.05	119	7507m	0.531	ug/l	
89) n-Butylbenzene	12.37	91	23495m	1.939	ug/l	
95) Naphthalene	13.82	128	25936	1.758	ug/l #	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_X\DATA\VX061920\
Data File : VX016894.D
Acq On : 19 Jun 2020 20:47
Operator : JC/SP
Sample : L3066-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-5DL

Manual Integrations
APPROVED

MMDadoda
6/22/2020 12:23:19 PM

Quant Time: Jul 03 06:23:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

