

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016906.D
 Acq On : 20 Jun 2020 04:22
 Operator : JC/SP
 Sample : L3063-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 39 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-4

Quant Time: Jun 22 07:15:48 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	323848	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	469707	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	571547	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	279477	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	162499	41.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	83.44%	
35) Dibromofluoromethane	5.47	113	170042	56.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	113.18%	
50) Toluene-d8	8.70	98	686437	57.10	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.20%	
62) 4-Bromofluorobenzene	11.12	95	289485	62.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	125.70%	
Target Compounds						
16) Acetone	2.42	43	5943	4.203	ug/l	93
17) Carbon Disulfide	2.55	76	2902	0.324	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016906.D
Acq On : 20 Jun 2020 04:22
Operator : JC/SP
Sample : L3063-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 39 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-4

Quant Time: Jun 22 07:15:48 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

