

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX061920\
 Data File : VX016911.D
 Acq On : 20 Jun 2020 06:16
 Operator : JC/SP
 Sample : L3063-10
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-10

Quant Time: Jun 22 07:34:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 19 12:07:48 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	328531	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	478781	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	477602	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	302431	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	178439	45.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	90.32%	
35) Dibromofluoromethane	5.46	113	166589	54.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.78%	
50) Toluene-d8	8.70	98	567890	46.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.68%	
62) 4-Bromofluorobenzene	11.12	95	263917	56.21	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.42%	
Target Compounds						
8) Diethyl Ether	2.17	74	1951	0.778	ug/l	95
16) Acetone	2.42	43	13315	9.283	ug/l	97
17) Carbon Disulfide	2.55	76	2179	0.240	ug/l #	94
20) Methylene Chloride	2.84	84	3690	1.073	ug/l #	89
40) Benzene	6.11	78	3646	0.275	ug/l	99
52) Toluene	8.77	92	2737	0.311	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX061920\
Data File : VX016911.D
Acq On : 20 Jun 2020 06:16
Operator : JC/SP
Sample : L3063-10
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-10

Quant Time: Jun 22 07:34:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X061820W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 19 12:07:48 2020
Response via : Initial Calibration

