

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
 Data File : VX036204.D
 Acq On : 19 Jun 2023 14:30
 Operator : JC/MD
 Sample : 03251-12RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANKRE

Quant Time: Jun 20 03:42:40 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jun 16 16:28:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	135060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	275668	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	258922	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	112844	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	164637	66.705	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 133.400%#	
35) Dibromofluoromethane	5.385	113	104044	55.658	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.320%	
50) Toluene-d8	8.647	98	347210	50.777	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.560%	
62) 4-Bromofluorobenzene	11.079	95	138083	50.026	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 100.060%	
Target Compounds						
16) Acetone	2.386	43	1251	1.220	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX061923\
Data File : VX036204.D
Acq On : 19 Jun 2023 14:30
Operator : JC/MD
Sample : 03251-12RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANKRE

Quant Time: Jun 20 03:42:40 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X061623W.M
Quant Title : SW846 8260
QLast Update : Fri Jun 16 16:28:40 2023
Response via : Initial Calibration

